

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: WALS01  
 Lab Code: ACE SDG No.: Q2487  
 Instrument ID: MSVOA\_N Calibration Date/Time: 07/03/2025 10:50  
 Lab File ID: VN087260.D Init. Calib. Date(s): 06/30/2025 06/30/2025  
 Heated Purge: (Y/N) N Init. Calib. Time(s): 12:28 14:51  
 GC Column: RXI-624 ID: 0.25 (mm)

| COMPOUND    | Amount   | Conc.  | %Dev     | MAX%D |
|-------------|----------|--------|----------|-------|
| 1,4-Dioxane | 1000.000 | 936.93 | 6.306598 | 20    |

All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.