

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>WALS01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2487</u>
Lab File ID:	<u>VN087260.D</u>	Date Analyzed:	<u>07/03/2025</u>
Instrument ID:	<u>MSVOA_N</u>	Time Analyzed:	<u>10:50</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	212674	13.788				
UPPER LIMIT	425348	14.288				
LOWER LIMIT	106337	13.288				
EPA SAMPLE NO.						
G4 (1.5)	157941	13.79				
G4 (10)	206244	13.79				
G3 (9)	175707	13.79				
G3 (3)	144966	13.79				
G2 (2.5)	214279	13.79				
G2 (9)	170569	13.79				
G1 (10)	138860	13.79				
G1 (4.5)	173051	13.79				
VN0703WBL01	171520	13.79				
VN0703WBS01	190045	13.79				
VN0703WBSD01	180696	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.