

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2487</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Walsh Construction Company II, LLC</u>	Datafile :	<u>VN087264.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0703WBSD01	Vinyl chloride	20	19.7	ug/L	99	5		65	117	19
	1,1-Dichloroethene	20	18.8	ug/L	94	1		74	110	20
	2-Butanone	100	98.2	ug/L	98	10		65	122	26
	Carbon Tetrachloride	20	18.1	ug/L	91	1		77	113	15
	Chloroform	20	18.8	ug/L	94	5		79	113	20
	Benzene	20	17.5	ug/L	88	2		82	109	15
	1,2-Dichloroethane	20	19.5	ug/L	98	12		80	115	20
	Trichloroethene	20	17.3	ug/L	86	2		77	113	15
	Tetrachloroethene	20	17.9	ug/L	90	1		67	123	15
	Chlorobenzene	20	17.9	ug/L	90	1		82	109	15