

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: WALS01
 Lab Code: ACE SDG NO.: Q2487
 Lab File ID: VN087260.D Date Analyzed: 07/03/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:50
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	241822	8.23	395302	9.11	360882	11.87
UPPER LIMIT	483644	8.729	790604	9.606	721764	12.365
LOWER LIMIT	120911	7.729	197651	8.606	180441	11.365
EPA SAMPLE NO.						
G4 (1.5)	179908	8.24	349516	9.11	347283	11.87
G4 (10)	180816	8.24	355191	9.11	356558	11.87
G3 (9)	180750	8.23	348767	9.11	343098	11.87
VN0703WBL01	197279	8.23	379432	9.11	350668	11.87
VN0703WBS01	227972	8.23	381904	9.11	348191	11.87
VN0703WBSD01	210099	8.23	355057	9.11	330452	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.