

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: WALS01
Lab Code: ACE SDG NO.: Q2487
Lab File ID: VN087282.D Date Analyzed: 07/07/2025
Instrument ID: MSVOA_N Time Analyzed: 09:48
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	253960	8.23	391455	9.11	371947	11.87
UPPER LIMIT	507920	8.73	782910	9.606	743894	12.365
LOWER LIMIT	126980	7.73	195728	8.606	185974	11.365
EPA SAMPLE NO.						
G3 (3)	178305	8.23	333547	9.11	319292	11.87
G2 (2.5)	229669	8.23	438833	9.11	427268	11.87
G2 (9)	193002	8.23	368733	9.11	358573	11.87
G1 (10)	180189	8.23	346519	9.11	337646	11.87
G1 (4.5)	183726	8.24	344094	9.11	347071	11.87
VN0707WBL01	183454	8.23	339704	9.11	317463	11.87
VN0707WBS02	214445	8.23	354114	9.11	330344	11.87

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.