

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: WALS01
 Lab Code: ACE SDG NO.: Q2487
 Lab File ID: VN087260.D Date Analyzed: 07/03/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:50
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	212674	13.788				
UPPER LIMIT	425348	14.288				
LOWER LIMIT	106337	13.288				
EPA SAMPLE NO.						
G4 (1.5)	162292	13.79				
G4 (10)	167009	13.79				
G3 (9)	162048	13.79				
VN0703WBL01	171520	13.79				
VN0703WBS01	190045	13.79				
VN0703WBSD01	180696	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.