

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: WALS01
 Lab Code: ACE SDG NO.: Q2487
 Lab File ID: VN087282.D Date Analyzed: 07/07/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:48
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	217810	13.788				
UPPER LIMIT	435620	14.288				
LOWER LIMIT	108905	13.288				
EPA SAMPLE NO.						
G3 (3)	147659	13.79				
G2 (2.5)	202126	13.79				
G2 (9)	170388	13.79				
G1 (10)	160324	13.79				
G1 (4.5)	162085	13.79				
VN0707WBL01	155374	13.79				
VN0707WBS02	176576	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.