

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070225\
 Data File : VY022903.D
 Acq On : 02 Jul 2025 10:37
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 03 02:17:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	97	0.00
2 T	Dichlorodifluoromethane	0.428	0.411	4.0	94	0.00
3 P	Chloromethane	0.816	0.897	-9.9	110	0.00
4 C	Vinyl Chloride	1.020	1.044	-2.4#	97	0.00
5 T	Bromomethane	0.802	0.808	-0.7	102	0.00
6 T	Chloroethane	0.686	0.718	-4.7	101	0.00
7 T	Trichlorofluoromethane	1.129	1.131	-0.2	94	0.00
8 T	Diethyl Ether	0.279	0.290	-3.9	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.516	0.529	-2.5	100	0.00
10 T	Methyl Iodide	0.562	0.592	-5.3	92	0.00
11 T	Tert butyl alcohol	0.037	0.034	8.1	87	0.00
12 CM	1,1-Dichloroethene	0.506	0.524	-3.6#	99	0.00
13 T	Acrolein	0.050	0.026	48.0#	52	0.00
14 T	Allyl chloride	0.778	0.858	-10.3	105	0.00
15 T	Acrylonitrile	0.116	0.125	-7.8	101	0.00
16 T	Acetone	0.105	0.130	-23.8	133	0.00
17 T	Carbon Disulfide	1.635	1.666	-1.9	97	0.00
18 T	Methyl Acetate	0.349	0.342	2.0	95	0.00
19 T	Methyl tert-butyl Ether	1.378	1.459	-5.9	99	0.00
20 T	Methylene Chloride	0.666	0.654	1.8	108	0.00
21 T	trans-1,2-Dichloroethene	0.578	0.601	-4.0	99	0.00
22 T	Diisopropyl ether	1.729	1.897	-9.7	103	0.00
23 T	Vinyl Acetate	0.955	1.075	-12.6	103	0.00
24 P	1,1-Dichloroethane	1.044	1.127	-8.0	102	0.00
25 T	2-Butanone	0.154	0.175	-13.6	111	0.00
26 T	2,2-Dichloropropane	0.875	0.968	-10.6	106	0.00
27 T	cis-1,2-Dichloroethene	0.672	0.696	-3.6	99	0.00
28 T	Bromochloromethane	0.439	0.455	-3.6	96	0.00
29 T	Tetrahydrofuran	0.097	0.105	-8.2	100	0.00
30 C	Chloroform	1.076	1.121	-4.2#	99	0.00
31 T	Cyclohexane	0.959	1.013	-5.6	104	0.00
32 T	1,1,1-Trichloroethane	0.929	0.989	-6.5	101	0.00
33 S	1,2-Dichloroethane-d4	0.558	0.550	1.4	96	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	0.00
35 S	Dibromofluoromethane	0.304	0.303	0.3	95	0.00
36 T	1,1-Dichloropropene	0.461	0.495	-7.4	101	0.00
37 T	Ethyl Acetate	0.199	0.220	-10.6	102	0.00
38 T	Carbon Tetrachloride	0.486	0.518	-6.6	101	0.00
39 T	Methylcyclohexane	0.596	0.656	-10.1	101	0.00
40 TM	Benzene	1.417	1.521	-7.3	99	0.00
41 T	Methacrylonitrile	0.122	0.119	2.5	95	0.00
42 TM	1,2-Dichloroethane	0.388	0.415	-7.0	99	0.00
43 T	Isopropyl Acetate	0.413	0.453	-9.7	99	0.00
44 TM	Trichloroethene	0.356	0.379	-6.5	97	0.00
45 C	1,2-Dichloropropane	0.332	0.363	-9.3#	102	0.00
46 T	Dibromomethane	0.188	0.194	-3.2	97	0.00
47 T	Bromodichloromethane	0.485	0.513	-5.8	98	0.00
48 T	Methyl methacrylate	0.199	0.222	-11.6	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070225\
 Data File : VY022903.D
 Acq On : 02 Jul 2025 10:37
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 03 02:17:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	88	0.00
50 S	Toluene-d8	1.207	1.217	-0.8	96	0.00
51 T	4-Methyl-2-Pentanone	0.210	0.238	-13.3	101	0.00
52 CM	Toluene	0.894	0.967	-8.2#	100	0.00
53 T	t-1,3-Dichloropropene	0.437	0.480	-9.8	102	0.00
54 T	cis-1,3-Dichloropropene	0.506	0.551	-8.9	101	0.00
55 T	1,1,2-Trichloroethane	0.244	0.260	-6.6	98	0.00
56 T	Ethyl methacrylate	0.328	0.365	-11.3	99	0.00
57 T	1,3-Dichloropropane	0.425	0.457	-7.5	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.145	0.180	-24.1	102	0.00
59 T	2-Hexanone	0.143	0.165	-15.4	104	0.00
60 T	Dibromochloromethane	0.315	0.331	-5.1	96	0.00
61 T	1,2-Dibromoethane	0.228	0.239	-4.8	97	0.00
62 S	4-Bromofluorobenzene	0.388	0.389	-0.3	96	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
64 T	Tetrachloroethene	0.472	0.498	-5.5	91	0.00
65 PM	Chlorobenzene	1.099	1.178	-7.2	99	0.00
66 T	1,1,1,2-Tetrachloroethane	0.372	0.395	-6.2	97	0.00
67 C	Ethyl Benzene	1.930	2.164	-12.1#	101	0.00
68 T	m/p-Xylenes	0.746	0.834	-11.8	101	0.00
69 T	o-Xylene	0.703	0.783	-11.4	101	0.00
70 T	Styrene	1.178	1.313	-11.5	99	0.00
71 P	Bromoform	0.207	0.220	-6.3	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
73 T	Isopropylbenzene	3.698	4.114	-11.2	101	0.00
74 T	N-amyl acetate	0.830	0.962	-15.9	100	0.00
75 P	1,1,2,2-Tetrachloroethane	0.596	0.635	-6.5	104	0.00
76 T	1,2,3-Trichloropropane	0.510	0.501	1.8	92	0.00
77 T	Bromobenzene	0.838	0.894	-6.7	98	0.00
78 T	n-propylbenzene	4.465	5.033	-12.7	103	0.00
79 T	2-Chlorotoluene	2.522	2.788	-10.5	101	0.00
80 T	1,3,5-Trimethylbenzene	2.985	3.337	-11.8	102	0.00
81 T	trans-1,4-Dichloro-2-butene	0.202	0.228	-12.9	107	0.00
82 T	4-Chlorotoluene	2.650	2.881	-8.7	100	0.00
83 T	tert-Butylbenzene	2.633	2.974	-13.0	102	0.00
84 T	1,2,4-Trimethylbenzene	2.989	3.339	-11.7	100	0.00
85 T	sec-Butylbenzene	3.961	4.512	-13.9	103	0.00
86 T	p-Isopropyltoluene	3.296	3.729	-13.1	102	0.00
87 T	1,3-Dichlorobenzene	1.683	1.822	-8.3	99	0.00
88 T	1,4-Dichlorobenzene	1.672	1.793	-7.2	99	0.00
89 T	n-Butylbenzene	3.101	3.570	-15.1	104	0.00
90 T	Hexachloroethane	0.657	0.713	-8.5	99	0.00
91 T	1,2-Dichlorobenzene	1.483	1.577	-6.3	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.101	0.105	-4.0	95	0.00
93 T	1,2,4-Trichlorobenzene	0.838	0.880	-5.0	97	0.00
94 T	Hexachlorobutadiene	0.474	0.482	-1.7	97	0.00
95 T	Naphthalene	1.515	1.584	-4.6	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070225\
Data File : VY022903.D
Acq On : 02 Jul 2025 10:37
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jul 03 02:17:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 08:29:52 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.724	0.729	-0.7	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6