

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY062325

Review By	Mahesh Dadoda	Review On	6/24/2025 8:24:26 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	6/24/2025 8:29:32 AM		
SubDirectory	VY062325	HP Acquire Method	MSVOA_Y	HP Processing Method	82y062325s.m

STD. NAME	STD REF.#
Tune/Reschk	VP134461
Initial Calibration Stds	VP134462,VP134463,VP134464,VP134465,VP134466,VP134467
CCC	
Internal Standard/PEM	VP133934
ICV/I.BLK	VP134468
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022775.D	23 Jun 2025 10:17		SY/MD	Ok
2	VSTDIC005	VSTDIC005	VY022776.D	23 Jun 2025 13:38	LR- 58	SY/MD	Ok,M
3	VSTDIC010	VSTDIC010	VY022777.D	23 Jun 2025 14:00		SY/MD	Ok,M
4	VSTDIC020	VSTDIC020	VY022778.D	23 Jun 2025 14:23		SY/MD	Ok,M
5	VSTDIC050	VSTDIC050	VY022779.D	23 Jun 2025 14:46		SY/MD	Ok,M
6	VSTDIC100	VSTDIC100	VY022780.D	23 Jun 2025 15:08		SY/MD	Ok,M
7	VSTDIC150	VSTDIC150	VY022781.D	23 Jun 2025 15:31		SY/MD	Ok,M
8	VIBLK	VIBLK	VY022782.D	23 Jun 2025 15:54		SY/MD	Ok
9	VSTDICV050	ICVVY062325	VY022783.D	23 Jun 2025 16:17		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY070225

Review By	Maresh Dadoda	Review On	7/3/2025 3:08:26 PM
Supervise By	Semsettin Yesilyurt	Supervise On	7/3/2025 3:13:05 PM
SubDirectory	VY070225	HP Acquire Method	MSVOA_Y HP Processing Method 82y062325s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134605
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134608,VP134609 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022902.D	02 Jul 2025 08:51		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022903.D	02 Jul 2025 10:37	CCAL failed high for comp.#16	SY/MD	Ok,M
3	VY0702SBL01	VY0702SBL01	VY022904.D	02 Jul 2025 11:42		SY/MD	Ok
4	VY0702SBS01	VY0702SBS01	VY022905.D	02 Jul 2025 12:12		SY/MD	Ok,M
5	VY0702SBSD01	VY0702SBSD01	VY022906.D	02 Jul 2025 12:35		SY/MD	Ok,M
6	Q2469-03	WC-1-VOC	VY022907.D	02 Jul 2025 13:12	vial-A	SY/MD	Ok
7	Q2464-01	OR-3-063025	VY022908.D	02 Jul 2025 13:36	vial-A	SY/MD	Ok
8	Q2478-03	WC-1-VOC	VY022909.D	02 Jul 2025 13:59	vial-A	SY/MD	Ok
9	Q2475-01	SOIL-PILE	VY022910.D	02 Jul 2025 14:23	vial-A	SY/MD	Ok
10	Q2486-02	VOC	VY022911.D	02 Jul 2025 14:46	vial-A Internal Standard Fail	SY/MD	ReRun
11	Q2458-04	TP-67	VY022912.D	02 Jul 2025 15:09	vial-B	SY/MD	Ok
12	Q2458-06	TP-60	VY022913.D	02 Jul 2025 15:33	vial-A	SY/MD	Ok
13	Q2458-07	TP-62	VY022914.D	02 Jul 2025 15:56	vial-A	SY/MD	Ok
14	Q2458-08	TP-63	VY022915.D	02 Jul 2025 16:20	vial-A	SY/MD	Ok
15	Q2458-09	TP-59	VY022916.D	02 Jul 2025 16:43	vial-A CCC Failed High	SY/MD	ReRun
16	Q2473-01	PIT#1	VY022917.D	02 Jul 2025 17:07	vial-A	SY/MD	Ok
17	Q2473-02	PIT#2	VY022918.D	02 Jul 2025 17:30	vial-A	SY/MD	Ok
18	Q2473-03	PIT#3	VY022919.D	02 Jul 2025 17:53	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QCBatch ID # VY070225

Review By	Mahesh Dadoda	Review On	7/3/2025 3:08:26 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	7/3/2025 3:13:05 PM		
SubDirectory	VY070225	HP Acquire Method	MSVOA_Y	HP Processing Method	82y062325s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP134605			
CCC		VP134608,VP134609			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q2473-04	PIT#4	VY022920.D	02 Jul 2025 18:17	vial-A Surrogate fail;CCC Failed High	SY/MD	ReRun
20	Q2487-01	G4(1.5)	VY022921.D	02 Jul 2025 18:40	vial-A CCC failed high	SY/MD	ReRun
21	Q2487-02	G4(10)	VY022922.D	02 Jul 2025 19:04	vial-A CCC failed high	SY/MD	ReRun
22	Q2487-03	G3(9)	VY022923.D	02 Jul 2025 19:27	vial-A CCC failed high	SY/MD	ReRun
23	Q2487-04	G3(3)	VY022924.D	02 Jul 2025 19:50	vial-A Internal standard fail;Surrogate fail	SY/MD	ReRun
24	Q2487-05	G2(2.5)	VY022925.D	02 Jul 2025 20:14	vial-A Internal Standard Fail	SY/MD	ReRun
25	VIBLK	VIBLK	VY022926.D	02 Jul 2025 20:37		SY/MD	Ok
26	VSTDCCC050	VSTDCCC050EC	VY022927.D	02 Jul 2025 21:23	Out of Tune	SY/MD	Not Ok

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY070325

Review By	Mahesh Dadoda	Review On	7/7/2025 2:11:58 PM
Supervise By	Semsettin Yesilyurt	Supervise On	7/7/2025 2:15:30 PM
SubDirectory	VY070325	HP Acquire Method	MSVOA_Y HP Processing Method 82y062325s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134628
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134631,VP134632 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022928.D	03 Jul 2025 08:03		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022929.D	03 Jul 2025 09:26		SY/MD	Ok,M
3	VY0703SBL01	VY0703SBL01	VY022930.D	03 Jul 2025 09:57		SY/MD	Ok
4	VY0703SBS01	VY0703SBS01	VY022931.D	03 Jul 2025 10:29		SY/MD	Ok,M
5	VY0703SBSD01	VY0703SBSD01	VY022932.D	03 Jul 2025 10:52		SY/MD	Ok,M
6	Q2493-03	WC-11-VOC	VY022933.D	03 Jul 2025 11:29	vial-A	SY/MD	Ok
7	Q2491-01	EO-1-070225	VY022934.D	03 Jul 2025 11:53	vial-A	SY/MD	Ok
8	Q2486-02	VOC	VY022935.D	03 Jul 2025 12:16	vial-B	SY/MD	Ok
9	Q2473-04RE	PIT#4RE	VY022936.D	03 Jul 2025 12:39	vial-B Internal Standard Fail; Surrogate Fail	SY/MD	Confirms
10	Q2458-09	TP-59	VY022937.D	03 Jul 2025 13:03	vial-B	SY/MD	Ok
11	Q2487-01	G4(1.5)	VY022938.D	03 Jul 2025 13:26	vial-B	SY/MD	Ok
12	Q2487-02	G4(10)	VY022939.D	03 Jul 2025 13:50	vial-B	SY/MD	Ok
13	Q2487-03	G3(9)	VY022940.D	03 Jul 2025 14:13	vial-B	SY/MD	Ok
14	Q2487-04	G3(3)	VY022941.D	03 Jul 2025 14:36	vial-B	SY/MD	Ok
15	Q2487-05RE	G2(2.5)RE	VY022942.D	03 Jul 2025 15:00	vial-B Internal Standard Fail	SY/MD	Confirms
16	Q2487-06	G2(9)	VY022943.D	03 Jul 2025 15:23	vial-A	SY/MD	Ok
17	Q2487-07	G1(4.5)	VY022944.D	03 Jul 2025 15:47	vial-A Internal Standard Fail; Surrogate Fail	SY/MD	ReRun

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY070725

Review By	Mahesh Dadoda	Review On	7/8/2025 3:35:09 PM
Supervise By	Semsettin Yesilyurt	Supervise On	7/8/2025 3:38:12 PM
SubDirectory	VY070725	HP Acquire Method	MSVOA_Y HP Processing Method 82y062325s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134656
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134657,VP134658 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022945.D	07 Jul 2025 08:36		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022946.D	07 Jul 2025 09:08		SY/MD	Ok,M
3	VY0707SBL01	VY0707SBL01	VY022947.D	07 Jul 2025 09:44		SY/MD	Ok
4	VY0707SBS01	VY0707SBS01	VY022948.D	07 Jul 2025 10:17	BS Failed High for com.#20	SY/MD	Ok,M
5	VY0707SBSD01	VY0707SBSD01	VY022949.D	07 Jul 2025 10:40	BSD Failed High for com.#20	SY/MD	Ok,M
6	Q2507-03	SU-04-7.3-2025-VOC	VY022950.D	07 Jul 2025 11:18	vial-A Internal Standard Fail	SY/MD	ReRun
7	Q2513-01	HR-2-070325	VY022951.D	07 Jul 2025 11:41	vial-A	SY/MD	Ok
8	Q2513-03	HR-3-070325	VY022952.D	07 Jul 2025 12:04	vial-A	SY/MD	Ok
9	Q2504-02	VOC	VY022953.D	07 Jul 2025 12:28	vial-A Internal Standard Fail;BS,BSD Failed High	SY/MD	ReRun
10	Q2515-01	wc-1	VY022954.D	07 Jul 2025 12:51	vial-A	SY/MD	Ok,M
11	Q2487-07RE	G1(4.5)RE	VY022955.D	07 Jul 2025 13:15	vial-B Internal standard fail	SY/MD	Confirms
12	Q2487-08	G1(10)	VY022956.D	07 Jul 2025 13:38	vial-A	SY/MD	Ok
13	Q2510-02	#63025-A-VOC	VY022957.D	07 Jul 2025 14:02	vial-A	SY/MD	Ok,M
14	Q2484-01	TP-58	VY022958.D	07 Jul 2025 14:25	vial-A	SY/MD	Ok
15	Q2484-02	TP-57	VY022959.D	07 Jul 2025 14:48	vial-A	SY/MD	Ok
16	Q2484-03	TP-64	VY022960.D	07 Jul 2025 15:12	vial-A	SY/MD	Ok
17	Q2484-04	TP-107	VY022961.D	07 Jul 2025 15:35	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY070725

Review By	Mahesh Dadoda	Review On	7/8/2025 3:35:09 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	7/8/2025 3:38:12 PM		
SubDirectory	VY070725	HP Acquire Method	MSVOA_Y	HP Processing Method	82y062325s.m
STD. NAME	STD REF.#				
Tune/Reschk Initial Calibration Stds	VP134656				
CCC	VP134657,VP134658				
Internal Standard/PEM	VP133934				
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q2484-05	TP-106	VY022962.D	07 Jul 2025 15:59	vial-A Internal Standard Fail; Surrogate Fail	SY/MD	ReRun
19	Q2484-06	TP-104	VY022963.D	07 Jul 2025 16:22	vial-A	SY/MD	Ok
20	Q2514-01	TP-92	VY022964.D	07 Jul 2025 16:45	vial-A	SY/MD	Ok
21	Q2514-02	TP-93	VY022965.D	07 Jul 2025 17:09	vial-A	SY/MD	Ok
22	Q2480-01	GPX1	VY022966.D	07 Jul 2025 17:32	vial-A	SY/MD	Ok

M : Manual Integration