

Surrogate Summary

SDG No.: Q2487

Client: Walsh Construction Company II, LLC

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2487-01	G4(1.5)	1,2-Dichloroethane-d4	50	53.3	107		63	155
		Dibromofluoromethane	50	53.7	107		70	134
		Toluene-d8	50	50.9	102		74	123
		4-Bromofluorobenzene	50	55.4	111		17	146
Q2487-02	G4(10)	1,2-Dichloroethane-d4	50	52.6	105		63	155
		Dibromofluoromethane	50	52.0	104		70	134
		Toluene-d8	50	51.5	103		74	123
		4-Bromofluorobenzene	50	56.7	113		17	146
Q2487-03	G3(9)	1,2-Dichloroethane-d4	50	53.5	107		63	155
		Dibromofluoromethane	50	53.5	107		70	134
		Toluene-d8	50	50.6	101		74	123
		4-Bromofluorobenzene	50	54.7	109		17	146
Q2487-04	G3(3)	1,2-Dichloroethane-d4	50	49.1	98		63	155
		Dibromofluoromethane	50	50.8	101		70	134
		Toluene-d8	50	50.5	101		74	123
		4-Bromofluorobenzene	50	50.9	102		17	146
Q2487-05	G2(2.5)	1,2-Dichloroethane-d4	50	44.5	89		63	155
		Dibromofluoromethane	50	55.2	110		70	134
		Toluene-d8	50	46.0	92		74	123
		4-Bromofluorobenzene	50	22.9	46		17	146
Q2487-05RE	G2(2.5)RE	1,2-Dichloroethane-d4	50	46.2	92		63	155
		Dibromofluoromethane	50	52.9	106		70	134
		Toluene-d8	50	48.3	97		74	123
		4-Bromofluorobenzene	50	29.0	58		17	146
Q2487-06	G2(9)	1,2-Dichloroethane-d4	50	53.6	107		63	155
		Dibromofluoromethane	50	53.2	106		70	134
		Toluene-d8	50	51.2	102		74	123
		4-Bromofluorobenzene	50	55.4	111		17	146
Q2487-07	G1(4.5)	1,2-Dichloroethane-d4	50	30.9	62	*	63	155
		Dibromofluoromethane	50	44.5	89		70	134
		Toluene-d8	50	48.2	96		74	123
		4-Bromofluorobenzene	50	34.7	69		17	146
Q2487-07RE	G1(4.5)RE	1,2-Dichloroethane-d4	50	39.8	80		63	155
		Dibromofluoromethane	50	49.1	98		70	134
		Toluene-d8	50	48.3	97		74	123
		4-Bromofluorobenzene	50	41.9	84		17	146
Q2487-08	G1(10)	1,2-Dichloroethane-d4	50	51.6	103		63	155
		Dibromofluoromethane	50	53.6	107		70	134
		Toluene-d8	50	49.9	100		74	123
		4-Bromofluorobenzene	50	45.5	91		17	146
VY0702SBL01	VY0702SBL01	1,2-Dichloroethane-d4	50	49.3	99		63	155
		Dibromofluoromethane	50	50.9	102		70	134
		Toluene-d8	50	51.3	103		74	123
		4-Bromofluorobenzene	50	55.3	111		17	146
VY0702SBS01	VY0702SBS01	1,2-Dichloroethane-d4	50	49.8	100		63	155
		Dibromofluoromethane	50	52.0	104		70	134
		Toluene-d8	50	51.9	104		74	123
		4-Bromofluorobenzene	50	48.9	98		17	146
VY0702SBSD01	VY0702SBSD01	1,2-Dichloroethane-d4	50	48.9	98		63	155
		Dibromofluoromethane	50	49.5	99		70	134

Surrogate Summary

SDG No.: Q2487

Client: Walsh Construction Company II, LLC

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
VY0702SBSD01	VY0702SBSD01	Toluene-d8	50	49.2	98		74	123
		4-Bromofluorobenzene	50	47.2	94		17	146
VY0703SBL01	VY0703SBL01	1,2-Dichloroethane-d4	50	52.7	105		63	155
		Dibromofluoromethane	50	52.7	105		70	134
		Toluene-d8	50	52.1	104		74	123
		4-Bromofluorobenzene	50	55.3	111		17	146
VY0703SBS01	VY0703SBS01	1,2-Dichloroethane-d4	50	48.4	97		63	155
		Dibromofluoromethane	50	50.6	101		70	134
		Toluene-d8	50	51.1	102		74	123
		4-Bromofluorobenzene	50	47.0	94		17	146
VY0707SBL01	VY0707SBL01	1,2-Dichloroethane-d4	50	45.8	91		63	155
		Dibromofluoromethane	50	50.4	101		70	134
		Toluene-d8	50	50.9	102		74	123
		4-Bromofluorobenzene	50	53.3	107		17	146
VY0707SBS01	VY0707SBS01	1,2-Dichloroethane-d4	50	46.9	94		63	155
		Dibromofluoromethane	50	49.5	99		70	134
		Toluene-d8	50	50.4	101		74	123
		4-Bromofluorobenzene	50	47.0	94		17	146