

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: WALS01
 Lab Code: ACE SDG NO.: Q2487
 Lab File ID: VY022903.D Date Analyzed: 07/02/2025
 Instrument ID: MSVOA_Y Time Analyzed: 10:37
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	453239	7.71	750175	8.62	644334	11.41
UPPER LIMIT	906478	8.213	1500350	9.116	1288670	11.914
LOWER LIMIT	226620	7.213	375088	8.116	322167	10.914
EPA SAMPLE NO.						
G2 (2.5)	250233	7.71	421989	8.62	262343 *	11.41
VY0702SBL01	327303	7.71	608180	8.62	599654	11.42
VY0702SBS01	424569	7.71	714057	8.62	598160	11.41
VY0702SBSD01	405936	7.71	690466	8.62	580305	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.