

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: WALS01
Lab Code: ACE SDG NO.: Q2487
Lab File ID: VY022946.D Date Analyzed: 07/07/2025
Instrument ID: MSVOA_Y Time Analyzed: 09:08
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	469181	7.71	754142	8.62	648632	11.41
UPPER LIMIT	938362	8.207	1508280	9.116	1297260	11.914
LOWER LIMIT	234591	7.207	377071	8.116	324316	10.914
EPA SAMPLE NO.						
G1 (4.5) RE	143475 *	7.71	238103 *	8.62	198851 *	11.41
G1 (10)	283210	7.71	516435	8.62	471025	11.41
VY0707SBL01	359794	7.71	659995	8.62	632698	11.41
VY0707SBS01	438661	7.71	730129	8.61	612594	11.41

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.