

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>WALS01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2487</u>
Lab File ID:	<u>VY022903.D</u>	Date Analyzed:	<u>07/02/2025</u>
Instrument ID:	<u>MSVOA_Y</u>	Time Analyzed:	<u>10:37</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>Y</u>

	IS4 AREA #	RT #				
12 HOUR STD	321198	13.347				
UPPER LIMIT	642396	13.847				
LOWER LIMIT	160599	12.847				
EPA SAMPLE NO.						
G2 (2.5)	52246 *	13.35				
VY0702SBL01	251367	13.35				
VY0702SBS01	274634	13.35				
VY0702SBSD01	268831	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.