

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>WALS01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2487</u>
Lab File ID:	<u>VY022929.D</u>	Date Analyzed:	<u>07/03/2025</u>
Instrument ID:	<u>MSVOA_Y</u>	Time Analyzed:	<u>09:26</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>Y</u>

	IS4 AREA #	RT #				
12 HOUR STD	273118	13.347				
UPPER LIMIT	546236	13.847				
LOWER LIMIT	136559	12.847				
EPA SAMPLE NO.						
G4 (1.5)	211646	13.35				
G4 (10)	221137	13.35				
G3 (9)	196307	13.34				
G3 (3)	206599	13.35				
G2 (2.5) RE	72021 *	13.35				
G2 (9)	207222	13.35				
G1 (4.5)	104418 *	13.35				
VY0703SBL01	230762	13.35				
VY0703SBS01	252496	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.