

ANALYTICAL RESULTS SUMMARY

GENERAL CHEMISTRY
METALS
SEMI-VOLATILE ORGANICS

PROJECT NAME : NWIRP BETHPAGE 112G08005-WE13

TETRA TECH NUS, INC.

661 Andersen Drive

Suite 200

Pittsburgh, PA - 15220-2745

Phone No: 412-921-7090

ORDER ID : Q2488

ATTENTION : Ernie Wu



Laboratory Certification ID # 20012



1) Signature Page	3
2) Case Narrative	4
2.1) SVOC-SIMGroup1- Case Narrative	4
2.2) Metals-AES- Case Narrative	6
2.3) Genchem- Case Narrative	7
3) Qualifier Page	8
4) QA Checklist	10
5) SVOC-SIMGroup1 Data	11
6) Metals-AES Data	37
7) Genchem Data	74
8) Shipping Document	83
8.1) CHAIN OF CUSTODY	84
8.2) Lab Certificate	85

1
2
3
4
5
6
7
8

Cover Page

Order ID : Q2488

Project ID : NWIRP Bethpage 112G08005-WE13

Client : Tetra Tech NUS, Inc.

Lab Sample Number

Q2488-01
Q2488-02

Client Sample Number

RW8-SP100-20250701
RW8-SP303-20250701

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 7/16/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012



CASE NARRATIVE

Tetra Tech NUS, Inc.

Project Name: NWIRP Bethpage 112G08005-WE13

Project Manager : Ernie Wu

Order ID # Q2488

Test Name: SVOC-SIMGroup1

A. Number of Samples and Date of Receipt:

2 Water samples were received on 07/02/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Metals Group4, SVOC-SIMGroup1, TDS and TSS. This data package contains results for SVOC-SIMGroup1.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_N using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGAThe analysis of SVOC-SIMGroup1 was based on method 8270-Modified and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements .

The Tuning criteria met requirements.

E. Additional Comments:

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is)."

The Sample #RW8-SP100-20250701 have the concentration of target compound below Method detection limits, therefore it is not reported as Hit in Form1.

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

2
2.1

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Tetra Tech NUS, Inc.

Project Name: NWIRP Bethpage 112G08005-WE13

Project Manager : Ernie Wu

Order ID # Q2488

Test Name: Metals Group4

A. Number of Samples and Date of Receipt:

2 Water samples were received on 07/02/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Metals Group4, SVOC-SIMGroup1, TDS and TSS. This data package contains results for Metals Group4.

C. Analytical Techniques:

The analysis of Metals Group4 was based on method 6010D and digestion based on method 3010 (waters).

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all parameters.

The Duplicate analysis met criteria for all parameters.

The Matrix Spike analysis met criteria for all parameters.

The Matrix Spike Duplicate analysis met criteria for all parameters.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

E. Additional Comments:

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is).

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Tetra Tech NUS, Inc.

Project Name: NWIRP Bethpage 112G08005-WE13

Project Manager : Ernie Wu

Order ID # Q2488

Test Name: TDS,TSS

A. Number of Samples and Date of Receipt:

2 Water samples were received on 07/02/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Metals Group4, SVOC-SIMGroup1, TDS and TSS. This data package contains results for TDS,TSS.

C. Analytical Techniques:

The analysis of TDS was based on method SM2540 C and The analysis of TSS was based on method SM2540 D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all parameters.

The Duplicate analysis met criteria for all parameters.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

E. Additional Comments:

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is).

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following “ Results Qualifiers” are used:

J	Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).
U	Indicates the analyte was analyzed for, but not detected.
ND	Indicates the analyte was analyzed for, but not detected
E	Indicates the reported value is estimated because of the presence of interference
M	Indicates Duplicate injection precision not met.
N	Indicates the spiked sample recovery is not within control limits.
S	Indicates the reported value was determined by the Method of Standard Addition (MSA).
*	Indicates that the duplicate analysis is not within control limits.
+	Indicates the correlation coefficient for the MSA is less than 0.995.
D	Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.
M	Method qualifiers “P” for ICP instrument “PM” for ICP when Microwave Digestion is used “CV” for Manual Cold Vapor AA “AV” for automated Cold Vapor AA “CA” for MIDI-Distillation Spectrophotometric “AS” for Semi -Automated Spectrophotometric “C” for Manual Spectrophotometric “T” for Titrimetric “NR” for analyte not required to be analyzed
OR	Indicates the analyte’s concentration exceeds the calibrated range of the instrument for that specific analysis.
Q	Indicates the LCS did not meet the control limits requirements
H	Sample Analysis Out Of Hold Time

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q2488

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 07/16/2025

LAB CHRONICLE

OrderID:	Q2488	OrderDate:	7/2/2025 11:43:00 AM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage 112G08005-WE13
Contact:	Ernie Wu	Location:	A42

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2488-01	RW8-SP100-20250701	Water			07/01/25			07/02/25
			SVOC-SIMGroup1	8270-Modified		07/03/25	07/14/25	
Q2488-02	RW8-SP303-20250701	Water			07/01/25			07/02/25
			SVOC-SIMGroup1	8270-Modified		07/03/25	07/14/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q2488
Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID :				0.000					
			Total Svoc :		0.00				
			Total Concentration:		0.00				



SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	07/01/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	07/02/25
Client Sample ID:	RW8-SP100-20250701	SDG No.:	Q2488
Lab Sample ID:	Q2488-01	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037485.D	1	07/03/25 08:58	07/14/25 10:26	PB168720

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.32		30 - 150		80%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.37		30 - 150		93%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.34		55 - 111		86%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.34		53 - 106		85%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.53		58 - 132		132%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2370	7.732				
1146-65-2	Naphthalene-d8	5650	10.509				
15067-26-2	Acenaphthene-d10	2640	14.355				
1517-22-2	Phenanthrene-d10	4840	17.099				
1719-03-5	Chrysene-d12	3620	21.286				
1520-96-3	Perylene-d12	3160	23.522				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	07/01/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	07/02/25
Client Sample ID:	RW8-SP303-20250701	SDG No.:	Q2488
Lab Sample ID:	Q2488-02	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037486.D	1	07/03/25 08:58	07/14/25 11:02	PB168720

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.30		30 - 150		75%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.38		30 - 150		94%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.35		55 - 111		86%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.36		53 - 106		89%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.49		58 - 132		123%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1820	7.731				
1146-65-2	Naphthalene-d8	4230	10.509				
15067-26-2	Acenaphthene-d10	1940	14.355				
1517-22-2	Phenanthrene-d10	3580	17.099				
1719-03-5	Chrysene-d12	2750	21.286				
1520-96-3	Perylene-d12	2330	23.525				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: Q2488

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB168720BL	PB168720BL	2-Methylnaphthalene-d10	0.4	0.35	86		30	150
		Fluoranthene-d10	0.4	0.34	84		30	150
		Nitrobenzene-d5	0.4	0.37	92		55	111
		2-Fluorobiphenyl	0.4	0.41	102		53	106
		Terphenyl-d14	0.4	0.40	99		58	132
PB168720BS	PB168720BS	2-Methylnaphthalene-d10	0.4	0.35	86		30	150
		Fluoranthene-d10	0.4	0.31	77		30	150
		Nitrobenzene-d5	0.4	0.35	87		55	111
		2-Fluorobiphenyl	0.4	0.36	90		53	106
		Terphenyl-d14	0.4	0.36	90		58	132
PB168720BSD	PB168720BSD	2-Methylnaphthalene-d10	0.4	0.34	85		30	150
		Fluoranthene-d10	0.4	0.32	80		30	150
		Nitrobenzene-d5	0.4	0.37	91		55	111
		2-Fluorobiphenyl	0.4	0.36	89		53	106
		Terphenyl-d14	0.4	0.38	95		58	132
Q2488-01	RW8-SP100-20250701	2-Methylnaphthalene-d10	0.4	0.32	80		30	150
		Fluoranthene-d10	0.4	0.37	93		30	150
		Nitrobenzene-d5	0.4	0.34	86		55	111
		2-Fluorobiphenyl	0.4	0.34	85		53	106
		Terphenyl-d14	0.4	0.53	132		58	132
Q2488-02	RW8-SP303-20250701	2-Methylnaphthalene-d10	0.4	0.30	75		30	150
		Fluoranthene-d10	0.4	0.38	94		30	150
		Nitrobenzene-d5	0.4	0.35	86		55	111
		2-Fluorobiphenyl	0.4	0.36	89		53	106
		Terphenyl-d14	0.4	0.49	123		58	132

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2488

Analytical Method: 8270-Modified

Client: Tetra Tech NUS, Inc.

DataFile: BN037487.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB168720BS	1,4-Dioxane	0.4	0.33	ug/L	83				70	130	

A
B
C
D
E
F
G

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2488</u>	Analytical Method:	<u>8270-Modified</u>
Client:	<u>Tetra Tech NUS, Inc.</u>	DataFile:	<u>BN037488.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB168720BSD	1,4-Dioxane	0.4	0.30	ug/L	75	10			70	130	20

4B

SEMIVOLATILE METHOD BLANK SUMMARY

Client ID

PB168720BL

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG NO.: Q2488

Lab File ID: BN037475.D

Lab Sample ID: PB168720BL

Instrument ID: BNA_N

Date Extracted: 07/03/2025

Matrix: (soil/water) Water

Date Analyzed: 07/11/2025

Level: (low/med) LOW

Time Analyzed: 16:02

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB168720BS	PB168720BS	BN037487.D	07/14/2025
RW8-SP100-20250701	Q2488-01	BN037485.D	07/14/2025
RW8-SP303-20250701	Q2488-02	BN037486.D	07/14/2025
PB168720BSD	PB168720BSD	BN037488.D	07/14/2025

COMMENTS: _____

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN037465.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q2488
DFTPP Injection Date: 07/11/2025
DFTPP Injection Time: 08:36

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.2 (0.6) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.1 (0.3) 1
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
365	Greater than 1% of mass 198	3.9
441	Present, but less than mass 443	78.9
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	15 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN037467.D	07/11/2025	11:10
SSTDICC0.2	SSTDICC0.2	BN037468.D	07/11/2025	11:46
SSTDICCC0.4	SSTDICCC0.4	BN037469.D	07/11/2025	12:22
SSTDICC0.8	SSTDICC0.8	BN037470.D	07/11/2025	12:59
SSTDICC1.6	SSTDICC1.6	BN037471.D	07/11/2025	13:36
SSTDICC3.2	SSTDICC3.2	BN037472.D	07/11/2025	14:12
SSTDICC5.0	SSTDICC5.0	BN037473.D	07/11/2025	14:49
PB168720BL	PB168720BL	BN037475.D	07/11/2025	16:02
SSTDCCC0.4EC	SSTDCCC0.4	BN037481.D	07/11/2025	19:43

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN037482.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q2488
DFTPP Injection Date: 07/14/2025
DFTPP Injection Time: 08:31

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.3 (0.7) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
365	Greater than 1% of mass 198	4
441	Present, but less than mass 443	64.7
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	14.9 (20) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN037483.D	07/14/2025	09:10
RW8-SP100-20250701	Q2488-01	BN037485.D	07/14/2025	10:26
RW8-SP303-20250701	Q2488-02	BN037486.D	07/14/2025	11:02
PB168720BS	PB168720BS	BN037487.D	07/14/2025	11:39
PB168720BSD	PB168720BSD	BN037488.D	07/14/2025	12:15
SSTDCCC0.4EC	SSTDCCC0.4	BN037496.D	07/14/2025	18:00

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2488

Client ID : SSTDICCC0.4

Date Analyzed: 07/11/2025

Lab File ID: BN037469.D

Time Analyzed: 12:22

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1754	7.732	4116	10.51	1977	14.37
UPPER LIMIT	3508	8.232	8232	11.009	3954	14.866
LOWER LIMIT	877	7.232	2058	10.009	988.5	13.866
EPA SAMPLE NO.						
01 PB168720BL	1870	7.73	4442	10.51	1998	14.37

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2488

Client ID: SSTDICCC0.4

Date Analyzed: 07/11/2025

Lab File ID: BN037469.D

Time Analyzed: 12:22

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	3577	17.099	2684	21.286	2478	23.528
UPPER LIMIT	7154	17.599	5368	21.786	4956	24.028
LOWER LIMIT	1788.5	16.599	1342	20.786	1239	23.028
EPA SAMPLE NO.						
01 PB168720BL	3362	17.10	2365	21.29	2144	23.53

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2488

Client ID : SSTDCCC0.4

Date Analyzed: 07/14/2025

Lab File ID: BN037483.D

Time Analyzed: 09:10

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1935	7.732	4597	10.51	2166	14.36
UPPER LIMIT	3870	8.232	9194	11.009	4332	14.855
LOWER LIMIT	967.5	7.232	2298.5	10.009	1083	13.855
EPA SAMPLE NO.						
01 PB168720BS	1966	7.73	4534	10.51	2005	14.36
02 PB168720BSD	1906	7.73	4315	10.51	1930	14.36
03 RW8-SP100-20250701	2365	7.73	5646	10.51	2637	14.36
04 RW8-SP303-20250701	1815	7.73	4225	10.51	1935	14.36

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2488

Client ID: SSTDCCC0.4

Date Analyzed: 07/14/2025

Lab File ID: BN037483.D

Time Analyzed: 09:10

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	3930	17.099	2837	21.286	2676	23.522
UPPER LIMIT	7860	17.599	5674	21.786	5352	24.022
LOWER LIMIT	1965	16.599	1418.5	20.786	1338	23.022
EPA SAMPLE NO.						
01 PB168720BS	3383	17.10	2498	21.29	2162	23.53
02 PB168720BSD	3183	17.10	2258	21.29	1944	23.53
03 RW8-SP100-20250701	4838	17.10	3617	21.29	3162	23.52
04 RW8-SP303-20250701	3578	17.10	2748	21.29	2331	23.53

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	PB168720BL		SDG No.:	Q2488
Lab Sample ID:	PB168720BL		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037475.D	1	07/03/25 08:58	07/11/25 16:02	PB168720

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.35		30 - 150		86%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.34		30 - 150		84%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.37		55 - 111		92%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.41		53 - 106		102%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.40		58 - 132		99%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1870	7.731				
1146-65-2	Naphthalene-d8	4440	10.509				
15067-26-2	Acenaphthene-d10	2000	14.366				
1517-22-2	Phenanthrene-d10	3360	17.099				
1719-03-5	Chrysene-d12	2370	21.286				
1520-96-3	Perylene-d12	2140	23.534				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:		
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:		
Client Sample ID:	PB168720BS		SDG No.:	Q2488	
Lab Sample ID:	PB168720BS		Matrix:	Water	
Analytical Method:	SW8270ESIM		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037487.D	1	07/03/25 08:58	07/14/25 11:39	PB168720

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.33		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.35		30 - 150		86%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.31		30 - 150		77%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.35		55 - 111		87%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.36		53 - 106		90%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.36		58 - 132		90%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1970	7.732				
1146-65-2	Naphthalene-d8	4530	10.509				
15067-26-2	Acenaphthene-d10	2010	14.355				
1517-22-2	Phenanthrene-d10	3380	17.099				
1719-03-5	Chrysene-d12	2500	21.286				
1520-96-3	Perylene-d12	2160	23.525				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:		
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:		
Client Sample ID:	PB168720BSD		SDG No.:	Q2488	
Lab Sample ID:	PB168720BSD		Matrix:	Water	
Analytical Method:	SW8270ESIM		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037488.D	1	07/03/25 08:58	07/14/25 12:15	PB168720

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.30		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.34		30 - 150		85%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.32		30 - 150		80%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.37		55 - 111		91%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.36		53 - 106		89%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.38		58 - 132		95%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1910	7.732				
1146-65-2	Naphthalene-d8	4320	10.509				
15067-26-2	Acenaphthene-d10	1930	14.356				
1517-22-2	Phenanthrene-d10	3180	17.099				
1719-03-5	Chrysene-d12	2260	21.286				
1520-96-3	Perylene-d12	1940	23.525				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\
 Method File : 8270-SIM-BN071125.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jul 11 15:38:34 2025
 Response Via : Initial Calibration

Calibration Files

0.1 =BN037467.D 0.2 =BN037468.D 0.4 =BN037469.D 0.8 =BN037470.D 1.6 =BN037471.D 3.2 =BN037472.D 5 =BN037473.D

	Compound	0.1	0.2	0.4	0.8	1.6	3.2	5	Avg	%RSD

1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2)	1,4-Dioxane	0.370	0.385	0.376	0.396	0.383	0.344	0.376		4.69
3)	n-Nitrosodimet...	0.478	0.499	0.474	0.526	0.520	0.517	0.502		4.49
4) S	2-Fluorophenol	0.941	0.935	0.931	0.854	0.927	0.942	0.987	0.931	4.22
5) S	Phenol-d6	1.222	1.196	1.173	1.063	1.158	1.171	1.214	1.171	4.52
6)	bis(2-Chloroet...	1.010	0.981	0.970	0.935	1.004	0.985	0.976	0.980	2.51
7) I	Naphthalene-d8	-----ISTD-----								
8) S	Nitrobenzene-d5	0.279	0.277	0.276	0.266	0.296	0.306	0.333	0.290	7.95
9)	Naphthalene	1.063	1.050	1.052	1.015	1.096	1.109	1.140	1.075	3.93
10)	Hexachlorobuta...	0.263	0.251	0.257	0.247	0.266	0.261	0.262	0.258	2.67
11) SURR2	Methylnaphth...	0.513	0.509	0.511	0.490	0.531	0.563	0.689	0.544	12.51
12)	2-Methylnaphth...	0.645	0.640	0.646	0.625	0.691	0.718	0.746	0.673	6.79
13) I	Acenaphthene-d10	-----ISTD-----								
14) S	2,4,6-Tribromo...	0.119	0.115	0.122	0.121	0.138	0.153	0.182	0.136	18.05
15) S	2-Fluorobiphenyl	2.213	2.152	2.305	2.139	2.318	2.357	2.499	2.283	5.55
16)	Acenaphthylene	1.778	1.768	1.761	1.706	1.881	1.886	1.988	1.824	5.34
17)	Acenaphthene	1.207	1.167	1.184	1.138	1.260	1.270	1.340	1.224	5.72
18)	Fluorene	1.478	1.432	1.482	1.423	1.606	1.620	1.723	1.538	7.35
19) I	Phenanthrene-d10	-----ISTD-----								
20)	4,6-Dinitro-2-...	0.036	0.039	0.043	0.050	0.057		0.045		19.20
21)	4-Bromophenyl-...	0.246	0.238	0.236	0.234	0.253	0.260	0.279	0.250	6.54
22)	Hexachlorobenzene	0.339	0.340	0.340	0.329	0.353	0.351	0.358	0.344	2.97
23)	Atrazine	0.118	0.111	0.115	0.114	0.133	0.153		0.124	13.13
24)	Pentachlorophenol		0.105	0.101	0.100	0.119	0.131	0.156	0.119	18.40
25)	Phenanthrene	1.213	1.165	1.157	1.147	1.242	1.256	1.346	1.218	5.79
26)	Anthracene	1.019	1.035	1.014	1.024	1.131	1.185	1.286	1.099	9.61
27) SURR	Fluoranthene-d10	0.956	0.936	0.909	0.895	0.981	1.049	1.336	1.009	15.16
28)	Fluoranthene	1.252	1.235	1.220	1.232	1.389	1.455	1.572	1.337	10.32
29) I	Chrysene-d12	-----ISTD-----								
30)	Pyrene	1.674	1.630	1.594	1.511	1.605	1.595	1.648	1.608	3.23
31) S	Terphenyl-d14	0.849	0.822	0.821	0.778	0.836	0.834	0.903	0.835	4.49
32)	Benzo(a)anthra...	1.329	1.310	1.292	1.240	1.394	1.428	1.490	1.355	6.40
33)	Chrysene	1.522	1.487	1.464	1.413	1.473	1.464	1.542	1.481	2.85
34)	Bis(2-ethylhex...	0.451	0.451	0.419	0.439	0.467	0.537	0.461		8.81
35) I	Perylene-d12	-----ISTD-----								

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\

Method File : 8270-SIM-BN071125.M

36)	Indeno(1,2,3-c...	1.339	1.316	1.392	1.434	1.655	1.758	1.973	1.553	16.01
37)	Benzo(b)fluora...	1.382	1.423	1.411	1.459	1.668	1.674	1.871	1.556	11.86
38)	Benzo(k)fluora...	1.416	1.374	1.433	1.505	1.674	1.745	1.892	1.577	12.41
39) C	Benzo(a)pyrene	1.069	1.109	1.152	1.162	1.325	1.398	1.549	1.252	14.09
40)	Dibenzo(a,h)an...	1.069	1.039	1.105	1.152	1.342	1.442	1.617	1.252	17.51
41)	Benzo(g,h,i)pe...	1.187	1.236	1.266	1.281	1.442	1.507	1.683	1.372	13.03

(#) = Out of Range

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2488</u>
Instrument ID:	<u>BNA_N</u>	Calibration Date/Time:	<u>07/11/2025 19:43</u>
Lab File ID:	<u>BN037481.D</u>	Init. Calib. Date(s):	<u>07/11/2025 07/11/2025</u>
EPA Sample No.:	<u>SSTDCCC0.4EC</u>	Init. Calib. Time(s):	<u>11:10 14:49</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.544	0.419		-23.0	50.0
Fluoranthene-d10	1.009	0.757		-25.0	50.0
2-Fluorophenol	0.931	0.915		-1.7	50.0
Phenol-d6	1.171	1.144		-2.3	50.0
Nitrobenzene-d5	0.290	0.238		-17.9	50.0
2-Fluorobiphenyl	2.283	1.999		-12.4	50.0
2,4,6-Tribromophenol	0.136	0.133		-2.2	50.0
Terphenyl-d14	0.835	0.664		-20.5	50.0
1,4-Dioxane	0.376	0.317		-15.7	50.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: TETR06
 Lab Code: ACE SDG No.: Q2488
 Instrument ID: BNA_N Calibration Date/Time: 07/14/2025 09:10
 Lab File ID: BN037483.D Init. Calib. Date(s): 07/11/2025 07/11/2025
 EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 11:10 14:49
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.544	0.511		-6.1	20.0
Fluoranthene-d10	1.009	0.900		-10.8	20.0
2-Fluorophenol	0.931	0.920		-1.2	20.0
Phenol-d6	1.171	1.170		-0.1	20.0
Nitrobenzene-d5	0.290	0.263		-9.3	20.0
2-Fluorobiphenyl	2.283	2.364		3.5	20.0
2,4,6-Tribromophenol	0.136	0.112		-17.6	20.0
Terphenyl-d14	0.835	0.853		2.2	20.0
1,4-Dioxane	0.376	0.399		6.1	20.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: TETR06
 Lab Code: ACE SDG No.: Q2488
 Instrument ID: BNA_N Calibration Date/Time: 07/14/2025 18:00
 Lab File ID: BN037496.D Init. Calib. Date(s): 07/11/2025 07/11/2025
 EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 11:10 14:49
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.544	0.510		-6.3	50.0
Fluoranthene-d10	1.009	0.908		-10.0	50.0
2-Fluorophenol	0.931	0.998		7.2	50.0
Phenol-d6	1.171	1.250		6.7	50.0
Nitrobenzene-d5	0.290	0.288		-0.7	50.0
2-Fluorobiphenyl	2.283	2.093		-8.3	50.0
2,4,6-Tribromophenol	0.136	0.153		12.5	50.0
Terphenyl-d14	0.835	0.740		-11.4	50.0
1,4-Dioxane	0.376	0.407		8.2	50.0

All other compounds must meet a minimum RRF of 0.010.

LAB CHRONICLE

OrderID:	Q2488	OrderDate:	7/2/2025 11:43:00 AM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage 112G08005-WE13
Contact:	Ernie Wu	Location:	A42

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2488-01	RW8-SP100-20250701	Water			07/01/25			07/02/25
			Metals Group4	6010D		07/07/25	07/09/25	
Q2488-02	RW8-SP303-20250701	Water			07/01/25			07/02/25
			Metals Group4	6010D		07/07/25	07/09/25	

Hit Summary Sheet
SW-846

SDG No.:	Q2488	Order ID:	Q2488
Client:	Tetra Tech NUS, Inc.	Project ID:	NWIRP Bethpage 112G08005-WE13

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID : RW8-SP100-20250701									
Q2488-01	RW8-SP100-20250701	Water	Iron	1040		11.7	40.0	50.0	ug/L
Client ID : RW8-SP303-20250701									
Q2488-02	RW8-SP303-20250701	Water	Iron	109		11.7	40.0	50.0	ug/L



SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	07/01/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	07/02/25
Client Sample ID:	RW8-SP100-20250701	SDG No.:	Q2488
Lab Sample ID:	Q2488-01	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7439-89-6	Iron	1040		1	11.7	40.0	50.0	ug/L	07/07/25 10:15	07/09/25 15:10	6010D	SW3010

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Metals Group4			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	07/01/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	07/02/25
Client Sample ID:	RW8-SP303-20250701	SDG No.:	Q2488
Lab Sample ID:	Q2488-02	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7439-89-6	Iron	109		1	11.7	40.0	50.0	ug/L	07/07/25 10:15	07/09/25 16:52	6010D	SW3010

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Metals Group4			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits



METAL CALIBRATION DATA

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2488

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	3730	4000	93	90 - 110	P	07/09/2025	10:05	LB136407

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2488

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	117	100	117	80 - 120	P	07/09/2025	10:17	LB136407

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

Contract: TETR06

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

SDG No.: Q2488

Lab Code: ACE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	5010	5000	100	90 - 110	P	07/09/2025	10:48	LB136407
CCV02	Iron	4860	5000	97	90 - 110	P	07/09/2025	11:47	LB136407
CCV03	Iron	4930	5000	99	90 - 110	P	07/09/2025	12:54	LB136407
CCV04	Iron	5000	5000	100	90 - 110	P	07/09/2025	14:05	LB136407
CCV05	Iron	5360	5000	107	90 - 110	P	07/09/2025	15:27	LB136407
CCV06	Iron	5080	5000	102	90 - 110	P	07/09/2025	16:39	LB136407
CCV07	Iron	5070	5000	101	90 - 110	P	07/09/2025	17:50	LB136407
CCV08	Iron	5020	5000	100	90 - 110	P	07/09/2025	18:54	LB136407
CCV09	Iron	5050	5000	101	90 - 110	P	07/09/2025	19:48	LB136407
CCV10	Iron	5110	5000	102	90 - 110	P	07/09/2025	20:41	LB136407
CCV11	Iron	5070	5000	101	90 - 110	P	07/09/2025	21:12	LB136407

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2488

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	4100	4000	103	90 - 110	P	07/10/2025	18:39	LB136434

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2488

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	118	100	118	80 - 120	P	07/10/2025	18:51	LB136434

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2488

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	5140	5000	103	90 - 110	P	07/10/2025	19:43	LB136434
CCV02	Iron	5010	5000	100	90 - 110	P	07/10/2025	20:59	LB136434
CCV03	Iron	5140	5000	103	90 - 110	P	07/10/2025	22:32	LB136434
CCV04	Iron	4890	5000	98	90 - 110	P	07/10/2025	23:29	LB136434
CCV05	Iron	4950	5000	99	90 - 110	P	07/11/2025	00:47	LB136434
CCV06	Iron	4890	5000	98	90 - 110	P	07/11/2025	01:47	LB136434
CCV07	Iron	4630	5000	93	90 - 110	P	07/11/2025	03:17	LB136434
CCV08	Iron	4780	5000	96	90 - 110	P	07/11/2025	04:18	LB136434
CCV09	Iron	4810	5000	96	90 - 110	P	07/11/2025	05:19	LB136434
CCV10	Iron	4700	5000	94	90 - 110	P	07/11/2025	06:16	LB136434
CCV11	Iron	4700	5000	94	90 - 110	P	07/11/2025	06:43	LB136434



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tetra Tech NUS, Inc.

SDG No.: Q2488

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Iron	119	100	119	65 - 135	P	07/09/2025	10:26	LB136407
CRI01	Iron	116	100	116	65 - 135	P	07/10/2025	19:01	LB136434



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc. **SDG No.:** Q2488
Contract: TETR06 **Lab Code:** ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Iron	25.4	+/-50	J	80.0	100	P	07/09/2025	10:22	LB136407

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc. **SDG No.:** Q2488
Contract: TETR06 **Lab Code:** ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Iron	23.4	+/-50	U	80.0	100	P	07/09/2025	10:57	LB136407
CCB02	Iron	23.4	+/-50	U	80.0	100	P	07/09/2025	11:56	LB136407
CCB03	Iron	23.4	+/-50	U	80.0	100	P	07/09/2025	12:58	LB136407
CCB04	Iron	23.4	+/-50	U	80.0	100	P	07/09/2025	14:28	LB136407
CCB05	Iron	23.4	+/-50	U	80.0	100	P	07/09/2025	15:33	LB136407
CCB06	Iron	23.4	+/-50	U	80.0	100	P	07/09/2025	16:43	LB136407
CCB07	Iron	23.4	+/-50	U	80.0	100	P	07/09/2025	17:54	LB136407
CCB08	Iron	23.4	+/-50	U	80.0	100	P	07/09/2025	18:58	LB136407
CCB09	Iron	23.4	+/-50	U	80.0	100	P	07/09/2025	19:52	LB136407
CCB10	Iron	23.4	+/-50	U	80.0	100	P	07/09/2025	20:45	LB136407
CCB11	Iron	23.4	+/-50	U	80.0	100	P	07/09/2025	21:16	LB136407

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc. **SDG No.:** Q2488
Contract: TETR06 **Lab Code:** ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Iron	23.4	+/-50	U	80.0	100	P	07/10/2025	18:55	LB136434

A

B

C

D

E

F

G

H

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q2488

Contract: TETR06

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Iron	23.4	+/-50	U	80.0	100	P	07/10/2025	19:54	LB136434
CCB02	Iron	23.4	+/-50	U	80.0	100	P	07/10/2025	21:04	LB136434
CCB03	Iron	23.4	+/-50	U	80.0	100	P	07/10/2025	22:36	LB136434
CCB04	Iron	23.4	+/-50	U	80.0	100	P	07/10/2025	23:53	LB136434
CCB05	Iron	23.4	+/-50	U	80.0	100	P	07/11/2025	00:51	LB136434
CCB06	Iron	23.4	+/-50	U	80.0	100	P	07/11/2025	02:23	LB136434
CCB07	Iron	23.4	+/-50	U	80.0	100	P	07/11/2025	03:26	LB136434
CCB08	Iron	23.4	+/-50	U	80.0	100	P	07/11/2025	04:27	LB136434
CCB09	Iron	23.4	+/-50	U	80.0	100	P	07/11/2025	05:24	LB136434
CCB10	Iron	23.4	+/-50	U	80.0	100	P	07/11/2025	06:21	LB136434
CCB11	Iron	23.4	+/-50	U	80.0	100	P	07/11/2025	06:47	LB136434

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q2488

Instrument: P5

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168739BL		WATER		Batch Number:		PB168739		Prep Date:	07/07/2025	
	Iron	11.7	<25	U	40.0	50.0	P	07/11/2025	00:56	LB136434

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Tetra Tech NUS, Inc.</u>	SDG No.: <u>Q2488</u>
Contract: <u>TETR06</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P5</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Iron	96800	100000	97	85600	116500	07/09/2025	10:31	LB136407
ICSAB01	Iron	98200	99000	99	84400	114500	07/09/2025	10:35	LB136407
ICSA01	Iron	104000	100000	104	85600	116500	07/10/2025	19:11	LB136434
ICSAB01	Iron	98700	99000	100	84400	114500	07/10/2025	19:15	LB136434



METAL QC DATA

metals
- 5a -
MATRIX SPIKE SUMMARY

client:	Tetra Tech NUS, Inc.	level:	low	sdg no.:	Q2488
contract:	TETR06			lab code:	ACE
matrix:	Water	sample id:	Q2512-01	client id:	WATER TREATMENT DISCHARGEMS
Percent Solids for Sample:	NA	Spiked ID:	Q2512-01MS	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Iron	ug/L	87 - 115	1720		186		1500	102		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client:	Tetra Tech NUS, Inc.	level:	low	sdg no.:	Q2488
contract:	TETR06			lab code:	ACE
matrix:	Water	sample id:	Q2512-01	client id:	WATER TREATMENT DISCHARGEMSD
Percent Solids for Sample:	NA	Spiked ID:	Q2512-01MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Iron	ug/L	87 - 115	1760		186		1500	105		P

Metals
- 5b -

Client: Tetra Tech NUS, Inc.

SDG No.: Q2488

Contract: TETR06

Lab Code: ACE

Matrix:

Level: LOW

Client ID:

Sample ID:

Spiked ID:

Analyte	Units	Acceptance Limit %R	C	Sample Result	C	Spike Added	% Recovery	Qual	M
---------	-------	------------------------	---	------------------	---	----------------	---------------	------	---

.....

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Tetra Tech NUS, Inc.</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q2488</u>
Contract:	<u>TETR06</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q2512-01</u>	Client ID:	<u>WATER TREATMENT DISCHARGEDU</u>
Percent Solids for Sample:	NA	Duplicate ID	Q2512-01DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Iron	ug/L	20	186		177		5		P

“A control limit of $\pm 20\%$ RPD for each matrix applies for sample values greater than 10 times Detection Limit”

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Tetra Tech NUS, Inc.</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q2488</u>
Contract:	<u>TETR06</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q2512-01MS</u>	Client ID:	<u>WATER TREATMENT DISCHARGEMS</u>
Percent Solids for Sample:	NA	Duplicate ID	Q2512-01MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Iron	ug/L	20	1720		1760		2		P

“A control limit of $\pm 20\%$ RPD for each matrix applies for sample values greater than 10 times Detection Limit”

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client:	<u>Tetra Tech NUS, Inc.</u>	SDG No.:	<u>Q2488</u>
Contract:	<u>TETR06</u>	Lab Code:	<u>ACE</u>

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168739BS Iron	ug/L	1500	1680		112	87 - 115	P

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

WATER TREATMENT DISCHARGEL

Lab Name: Alliance **Contract:** TETR06
Lab Code: ACE **Lb No.:** lb136434 **Lab Sample ID :** Q2512-01L **SDG No.:** Q2488
Matrix (soil/water): Water **Level (low/med):** LOW
Concentration Units: ug/L

Analyte	Initial Sample Result (I) C	Serial Dilution Result (S) C	% Differ- ence	Q	M
Iron	186	202 J	9		P



METAL PREPARATION & INSTRUMENT DATA

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Tetra Tech NUS, Inc.

SDG No.: Q2488

Contract: TETR06

Lab Code: ACE

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		Al	Ca	Fe	Mg	Ag
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Tetra Tech NUS, Inc.

SDG No.: Q2488

Contract: TETR06

Lab Code: ACE

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		As	Ba	Be	Cd	Co
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Tetra Tech NUS, Inc.

SDG No.: Q2488

Contract: TETR06

Lab Code: ACE

Instrument ID:

Date:

Interement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		Cr	Cu	K	Mn	Mo
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Tetra Tech NUS, Inc.

SDG No.: Q2488

Contract: TETR06

Lab Code: ACE

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		Na	Ni	Pb	Sb	Se
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Tetra Tech NUS, Inc.

SDG No.: Q2488

Contract: TETR06

Lab Code: ACE

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Sn	Ti	Tl	V	Zn
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000



METAL PREPARATION & ANALYICAL SUMMARY

Metals
- 13 -

SAMPLE PREPARATION SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q2488

Contract: TETR06

Lab Code: ACE

Method:

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(mL)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB168739							
PB168739BL	PB168739BL	MB	WATER	07/07/2025	50.0	25.0	
PB168739BS	PB168739BS	LCS	WATER	07/07/2025	50.0	25.0	
Q2488-01	RW8-SP100-20250701	SAM	WATER	07/07/2025	50.0	25.0	
Q2488-02	RW8-SP303-20250701	SAM	WATER	07/07/2025	50.0	25.0	
Q2512-01DUP	WATER TREATMENT DISCHARGEDUP	DUP	WATER	07/07/2025	50.0	25.0	
Q2512-01MS	WATER TREATMENT DISCHARGEMS	MS	WATER	07/07/2025	50.0	25.0	
Q2512-01MSD	WATER TREATMENT DISCHARGEMSD	MSD	WATER	07/07/2025	50.0	25.0	

metals
- 14 -
ANALYSIS RUN LOG

Client: Tetra Tech NUS, Inc.

Contract: TETR06

Lab code: ACE

Sdg no.: Q2488

Instrument id number:

Method:

Run number: LB136407

Start date: 07/09/2025

End date: 07/09/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	0939	Fe
S1	S1	1	0944	Fe
S2	S2	1	0948	Fe
S3	S3	1	0952	Fe
S4	S4	1	0957	Fe
S5	S5	1	1001	Fe
ICV01	ICV01	1	1005	Fe
LLICV01	LLICV01	1	1017	Fe
ICB01	ICB01	1	1022	Fe
CRI01	CRI01	1	1026	Fe
ICSA01	ICSA01	1	1031	Fe
ICSAB01	ICSAB01	1	1035	Fe
CCV01	CCV01	1	1048	Fe
CCB01	CCB01	1	1057	Fe
CCV02	CCV02	1	1147	Fe
CCB02	CCB02	1	1156	Fe
CCV03	CCV03	1	1254	Fe
CCB03	CCB03	1	1258	Fe
CCV04	CCV04	1	1405	Fe
CCB04	CCB04	1	1428	Fe
Q2488-01	RW8-SP100-20250701	1	1510	Fe
CCV05	CCV05	1	1527	Fe
CCB05	CCB05	1	1533	Fe
CCV06	CCV06	1	1639	Fe
CCB06	CCB06	1	1643	Fe
Q2488-02	RW8-SP303-20250701	1	1652	Fe
CCV07	CCV07	1	1750	Fe
CCB07	CCB07	1	1754	Fe
CCV08	CCV08	1	1854	Fe
CCB08	CCB08	1	1858	Fe
CCV09	CCV09	1	1948	Fe
CCB09	CCB09	1	1952	Fe
CCV10	CCV10	1	2041	Fe
CCB10	CCB10	1	2045	Fe
CCV11	CCV11	1	2112	Fe
CCB11	CCB11	1	2116	Fe

metals
- 14 -
ANALYSIS RUN LOG

Client: Tetra Tech NUS, Inc.

Contract: TETR06

Lab code: ACE

Sdg no.: Q2488

Instrument id number:

Method:

Run number: LB136434

Start date: 07/10/2025

End date: 07/11/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1646	Fe
S1	S1	1	1650	Fe
S2	S2	1	1655	Fe
S3	S3	1	1659	Fe
S4	S4	1	1703	Fe
S5	S5	1	1707	Fe
ICV01	ICV01	1	1839	Fe
LLICV01	LLICV01	1	1851	Fe
ICB01	ICB01	1	1855	Fe
CRI01	CRI01	1	1901	Fe
ICSA01	ICSA01	1	1911	Fe
ICSAB01	ICSAB01	1	1915	Fe
CCV01	CCV01	1	1943	Fe
CCB01	CCB01	1	1954	Fe
CCV02	CCV02	1	2059	Fe
CCB02	CCB02	1	2104	Fe
Q2512-01DUP	WATER TREATMENT DISCHA	1	2127	Fe
Q2512-01L	WATER TREATMENT DISCHA	5	2136	Fe
Q2512-01MS	WATER TREATMENT DISCHA	1	2150	Fe
Q2512-01MSD	WATER TREATMENT DISCHA	1	2159	Fe
CCV03	CCV03	1	2232	Fe
CCB03	CCB03	1	2236	Fe
CCV04	CCV04	1	2329	Fe
CCB04	CCB04	1	2353	Fe
CCV05	CCV05	1	0047	Fe
CCB05	CCB05	1	0051	Fe
PB168739BL	PB168739BL	1	0056	Fe
PB168739BS	PB168739BS	1	0100	Fe
CCV06	CCV06	1	0147	Fe
CCB06	CCB06	1	0223	Fe
CCV07	CCV07	1	0317	Fe
CCB07	CCB07	1	0326	Fe
CCV08	CCV08	1	0418	Fe
CCB08	CCB08	1	0427	Fe
CCV09	CCV09	1	0519	Fe
CCB09	CCB09	1	0524	Fe
CCV10	CCV10	1	0616	Fe
CCB10	CCB10	1	0621	Fe
CCV11	CCV11	1	0643	Fe
CCB11	CCB11	1	0647	Fe

LAB CHRONICLE

OrderID:	Q2488	OrderDate:	7/2/2025 11:43:00 AM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage 112G08005-WE13
Contact:	Ernie Wu	Location:	A42

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2488-02	RW8-SP303-20250701	WATER			07/01/25 13:13			07/02/25
			TDS	SM2540 C			07/02/25 16:30	
			TSS	SM2540 D			07/03/25 09:30	



SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	07/01/25 13:13
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	07/02/25
Client Sample ID:	RW8-SP303-20250701	SDG No.:	Q2488
Lab Sample ID:	Q2488-02	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
TDS	64.0		1	1.00	10.0	10.0	mg/L		07/02/25 16:30	SM 2540 C-20
TSS	4.00	U	1	1.00	4.00	4.00	mg/L		07/03/25 09:30	SM 2540 D-20

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements
H = Sample Analysis Out Of Hold Time

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N = Spiked sample recovery not within control limits



QC RESULT SUMMARY

Preparation Blank Summary

Client: Tetra Tech NUS, Inc.

SDG No.: Q2488

Project: NWIRP Bethpage 112G08005-WE13

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB136363BL							
TDS	mg/L	< 5.0000	5.0000	U	1.0	10	07/02/2025
Sample ID: LB136376BL							
TSS	mg/L	1	2.0000	J	1	4	07/03/2025

Duplicate Sample Summary

Client:	Tetra Tech NUS, Inc.	SDG No.:	Q2488
Project:	NWIRP Bethpage 112G08005-WE13	Sample ID:	Q2488-02
Client ID:	RW8-SP303-20250701DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
TDS	mg/L	+/-5	64.0		61.0		1	4.8		07/02/2025

Duplicate Sample Summary

Client:	Tetra Tech NUS, Inc.	SDG No.:	Q2488
Project:	NWIRP Bethpage 112G08005-WE13	Sample ID:	Q2499-02
Client ID:	CompDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
TSS	mg/L	+/-5	338		341		1	0.88		07/03/2025

Laboratory Control Sample Summary

Client:	Tetra Tech NUS, Inc.	SDG No.:	Q2488
Project:	NWIRP Bethpage 112G08005-WE13	Run No.:	LB136363

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136363BS							
TDS	mg/L	100	95.0		95	1	90-110	07/02/2025

Laboratory Control Sample Summary

Client:	Tetra Tech NUS, Inc.	SDG No.:	Q2488
Project:	NWIRP Bethpage 112G08005-WE13	Run No.:	LB136376

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136376BS							
TSS	mg/L	550	533		97	1	90-110	07/03/2025



SHIPPING DOCUMENTS

CHEMTECH CHAIN OF CUSTODY RECORD		284 Sheffield Street, Mountainside, NJ 07092 (908) 789-8900 Fax: (908) 78-8922 www.chemtech.net		Chemtech Project Number: Q 2488																					
				COC Number:																					
CLIENT INFORMATION		PROJECT INFORMATION		BILLING INFORMATION																					
COMPANY: Tetra Tech		PROJECT NAME: NWIRP Bethpage		BILL TO: PO#																					
ADDRESS: 4433 Corporation Ln, Suite 300		PROJECT #: 112G08005-WE13 LOCATION: RW8		ADDRESS:																					
CITY: Virginia Beach STATE: VA ZIP: 23462		PROJECT MANAGER: Ernie Wu		CITY: STATE: ZIP:																					
ATTENTION: Ernie Wu		E-MAIL: ernie.wu@tetratech.com		ATTENTION: PHONE:																					
PHONE: 757-466-4901 FAX: 757-461-4148		PHONE: 757-466-4901 FAX: 757-461-4148																							
DATA TURNAROUND INFORMATION		DATA DELIVERABLE INFORMATION		ANALYSIS																					
FAX: _____ 10 _____ DAYS* HARD COPY: _____ 10 _____ DAYS* EDD _____ 10 _____ DAYS* * TO BE APPROVED BY CHEMTECH STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS		☐ RESEULTS ONLY ☐ USEPA CLP ☐ RESULTS + QC ☐ New York State ASP "B" ☐ New Jersey REDUCED ☐ New York State ASP "A" ☐ New Jersey CLP ☐ Other _____ ☐ EDD Format _____		<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td style="width: 10%;">1,4-Dioxane SW846 8270</td> <td style="width: 10%;">Iron, Total</td> <td style="width: 10%;">TSS</td> <td style="width: 10%;">TDS</td> <td style="width: 10%;"> </td> <td style="width: 10%;"> </td> <td style="width: 10%;"> </td> <td style="width: 10%;"> </td> <td style="width: 10%;"> </td> <td style="width: 10%;"> </td> </tr> <tr> <td>1</td> <td>2</td> <td>3</td> <td>4</td> <td>5</td> <td>6</td> <td>7</td> <td>8</td> <td>9</td> <td> </td> </tr> </table>		1,4-Dioxane SW846 8270	Iron, Total	TSS	TDS							1	2	3	4	5	6	7	8	9	
1,4-Dioxane SW846 8270	Iron, Total	TSS	TDS																						
1	2	3	4	5	6	7	8	9																	
				PRESERVATIVES																					
				COMMENTS																					
CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE COMP GRAB	SAMPLE COLLECTION DATE TIME	# of Bottles																				
1.	RW8-SP100-20250701	GW		7/1/25 13:05	2																				
2.	RW8-SP303-20250701	GW		7/1/25 13:13	4																				
3.																									
4.																									
5.																									
6.																									
7.																									
8.																									
9.																									
10.																									
SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE PROSESSION INCLUDING COURIER DELIVERY																									
RELINQUISHED BY SAMPLER	DATE/TIME	RECEIVED BY	Conditions of bottles or coolers at receipt: ☐ Compliant ☐ Non Compliant ☐ Cooler Temp <u>2.1°C</u>																						
<i>[Signature]</i>	7/1/25 1430	<i>[Signature]</i>	MeOH extraction requires an additional 4oz. Jar for percent solid																						
RELINQUISHED BY	DATE/TIME	RECEIVED BY	Comments:																						
<i>[Signature]</i>	7/2/25 1000	<i>[Signature]</i>																							
RELINQUISHED BY	DATE/TIME	RECEIVED FOR LAB BY	SHIPPED VIA: CLIENT: ☐ Hand Delivered ☐ Overnight																						
<i>[Signature]</i>			CHEMTECH: ☐ Picked Up ☐ Overnight																						
3.			Shipment Complete ☐ YES ☐ NO																						
Page _____ of _____																									
WHITE - CHEMTECH COPY FOR RETURN TO CLIENT YELLOW - CHEMTECH COPY PINK - SAMPLER COPY																									

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488