

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2488

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	3730	4000	93	90 - 110	P	07/09/2025	10:05	LB136407



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.
Contract: TETR06
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

SDG No.: Q2488
Lab Code: ACE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	117	100	117	80 - 120	P	07/09/2025	10:17	LB136407

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2488

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	5010	5000	100	90 - 110	P	07/09/2025	10:48	LB136407
CCV02	Iron	4860	5000	97	90 - 110	P	07/09/2025	11:47	LB136407
CCV03	Iron	4930	5000	99	90 - 110	P	07/09/2025	12:54	LB136407
CCV04	Iron	5000	5000	100	90 - 110	P	07/09/2025	14:05	LB136407
CCV05	Iron	5360	5000	107	90 - 110	P	07/09/2025	15:27	LB136407
CCV06	Iron	5080	5000	102	90 - 110	P	07/09/2025	16:39	LB136407
CCV07	Iron	5070	5000	101	90 - 110	P	07/09/2025	17:50	LB136407
CCV08	Iron	5020	5000	100	90 - 110	P	07/09/2025	18:54	LB136407
CCV09	Iron	5050	5000	101	90 - 110	P	07/09/2025	19:48	LB136407
CCV10	Iron	5110	5000	102	90 - 110	P	07/09/2025	20:41	LB136407
CCV11	Iron	5070	5000	101	90 - 110	P	07/09/2025	21:12	LB136407

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2488

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	4100	4000	103	90 - 110	P	07/10/2025	18:39	LB136434

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2488

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	118	100	118	80 - 120	P	07/10/2025	18:51	LB136434

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2488

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	5140	5000	103	90 - 110	P	07/10/2025	19:43	LB136434
CCV02	Iron	5010	5000	100	90 - 110	P	07/10/2025	20:59	LB136434
CCV03	Iron	5140	5000	103	90 - 110	P	07/10/2025	22:32	LB136434
CCV04	Iron	4890	5000	98	90 - 110	P	07/10/2025	23:29	LB136434
CCV05	Iron	4950	5000	99	90 - 110	P	07/11/2025	00:47	LB136434
CCV06	Iron	4890	5000	98	90 - 110	P	07/11/2025	01:47	LB136434
CCV07	Iron	4630	5000	93	90 - 110	P	07/11/2025	03:17	LB136434
CCV08	Iron	4780	5000	96	90 - 110	P	07/11/2025	04:18	LB136434
CCV09	Iron	4810	5000	96	90 - 110	P	07/11/2025	05:19	LB136434
CCV10	Iron	4700	5000	94	90 - 110	P	07/11/2025	06:16	LB136434
CCV11	Iron	4700	5000	94	90 - 110	P	07/11/2025	06:43	LB136434