

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tetra Tech NUS, Inc.

SDG No.: Q2488

Contract: TETR06

Lab Code: ACE

ICS Source: EPA

Instrument ID: P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Iron	96800	100000	97	85600	116500	07/09/2025	10:31	LB136407
ICSAB01	Iron	98200	99000	99	84400	114500	07/09/2025	10:35	LB136407
ICSA01	Iron	104000	100000	104	85600	116500	07/10/2025	19:11	LB136434
ICSAB01	Iron	98700	99000	100	84400	114500	07/10/2025	19:15	LB136434