

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: TETR06
Lab Code: ACE SDG No.: Q2488
Instrument ID: BNA_N Calibration Date/Time: 07/11/2025 12:22
Lab File ID: BN037469.D Init. Calib. Date(s): _____
EPA Sample No.: SSTDICCC0.4 Init. Calib. Time(s): _____
GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.544	0.511		-6.1	20.0
Fluoranthene-d10	1.009	0.909		-9.9	20.0
2-Fluorophenol	0.931	0.931		0.0	20.0
Phenol-d6	1.171	1.173		0.2	20.0
Nitrobenzene-d5	0.290	0.276		-4.8	20.0
2-Fluorobiphenyl	2.283	2.305		1.0	20.0
2,4,6-Tribromophenol	0.136	0.122		-10.3	20.0
Terphenyl-d14	0.835	0.821		-1.7	20.0
1,4-Dioxane	0.376	0.385		2.4	20.0

All other compounds must meet a minimum RRF of 0.010.