

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN071125

Review By	Rahul	Review On	7/14/2025 11:11:52 AM
Supervise By	Jagrut	Supervise On	7/14/2025 2:29:11 PM
SubDirectory	BN071125	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method bn071125
STD. NAME	STD REF.#		
Tune/Reschk	SP6757		
Initial Calibration Stds	SP6842,SP6843,SP6844,SP6845,SP6846,SP6847,SP6848		
CCC	SP6846		
Internal Standard/PEM	SP6740,1ul/100ul sample		
ICV/I.BLK	SP6854		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN037465.D	11 Jul 2025 08:36	RC/JU	Ok
2	SSTDCCC0.4	BN037466.D	11 Jul 2025 09:51	RC/JU	Not Ok
3	SSTDICC0.1	BN037467.D	11 Jul 2025 11:10	RC/JU	Ok
4	SSTDICC0.2	BN037468.D	11 Jul 2025 11:46	RC/JU	Ok
5	SSTDICCC0.4	BN037469.D	11 Jul 2025 12:22	RC/JU	Ok
6	SSTDICC0.8	BN037470.D	11 Jul 2025 12:59	RC/JU	Ok
7	SSTDICC1.6	BN037471.D	11 Jul 2025 13:36	RC/JU	Ok
8	SSTDICC3.2	BN037472.D	11 Jul 2025 14:12	RC/JU	Ok
9	SSTDICC5.0	BN037473.D	11 Jul 2025 14:49	RC/JU	Ok
10	SSTDICV0.4	BN037474.D	11 Jul 2025 15:25	RC/JU	Ok
11	PB168720BL	BN037475.D	11 Jul 2025 16:02	RC/JU	Ok
12	Q2126-09	BN037476.D	11 Jul 2025 16:39	RC/JU	Ok,M
13	Q2126-09	BN037477.D	11 Jul 2025 17:16	RC/JU	Ok
14	Q2126-03	BN037478.D	11 Jul 2025 17:52	RC/JU	Ok
15	Q2126-03	BN037479.D	11 Jul 2025 18:29	RC/JU	Ok
16	PB168721BL	BN037480.D	11 Jul 2025 19:06	RC/JU	Ok
17	SSTDCCC0.4	BN037481.D	11 Jul 2025 19:43	RC/JU	Ok,M

M : Manual Integration

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN071425

Review By	Rahul	Review On	7/15/2025 1:04:59 PM
Supervise By	Jagrut	Supervise On	7/15/2025 2:43:18 PM
SubDirectory	BN071425	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method bn071125
STD. NAME	STD REF.#		
Tune/Reschk	SP6757		
Initial Calibration Stds	SP6842,SP6843,SP6844,SP6845,SP6846,SP6847,SP6848		
CCC	SP6846		
Internal Standard/PEM	SP6740,1ul/100ul sample		
ICV/I.BLK	SP6854		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN037482.D	14 Jul 2025 08:31	RC/JU	Ok
2	SSTDCCC0.4	BN037483.D	14 Jul 2025 09:10	RC/JU	Ok
3	PB168696BL	BN037484.D	14 Jul 2025 09:49	RC/JU	Ok
4	Q2488-01	BN037485.D	14 Jul 2025 10:26	RC/JU	Ok
5	Q2488-02	BN037486.D	14 Jul 2025 11:02	RC/JU	Ok
6	PB168720BS	BN037487.D	14 Jul 2025 11:39	RC/JU	Ok
7	PB168720BSD	BN037488.D	14 Jul 2025 12:15	RC/JU	Ok,M
8	Q2580-02	BN037489.D	14 Jul 2025 13:27	RC/JU	Not Ok
9	Q2580-02	BN037490.D	14 Jul 2025 14:21	RC/JU	Dilution
10	Q2580-02MS	BN037491.D	14 Jul 2025 14:57	RC/JU	Ok,M
11	Q2580-02MSD	BN037492.D	14 Jul 2025 15:34	RC/JU	Ok,M
12	Q2580-02DL	BN037493.D	14 Jul 2025 16:11	RC/JU	Ok,M
13	PB168831BS	BN037494.D	14 Jul 2025 16:47	RC/JU	Ok,M
14	PB168831BL	BN037495.D	14 Jul 2025 17:24	RC/JU	Ok
15	SSTDCCC0.4	BN037496.D	14 Jul 2025 18:00	RC/JU	Ok

M : Manual Integration