

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN071125

Review By	Rahul	Review On	7/14/2025 11:11:52 AM
Supervise By	Jagrut	Supervise On	7/14/2025 2:29:11 PM
SubDirectory	BN071125	HP Acquire Method	BNA_N, 8270_HP Processing Method bn071125

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6842,SP6843,SP6844,SP6845,SP6846,SP6847,SP6848
CCC	SP6846
Internal Standard/PEM	SP6740,1ul/100ul sample
ICV/I.BLK	SP6854
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN037465.D	11 Jul 2025 08:36		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN037466.D	11 Jul 2025 09:51	A Fresh Calibration is required.	RC/JU	Not Ok
3	SSTDICC0.1	SSTDICC0.1	BN037467.D	11 Jul 2025 11:10		RC/JU	Ok
4	SSTDICC0.2	SSTDICC0.2	BN037468.D	11 Jul 2025 11:46		RC/JU	Ok
5	SSTDICCC0.4	SSTDICCC0.4	BN037469.D	11 Jul 2025 12:22		RC/JU	Ok
6	SSTDICC0.8	SSTDICC0.8	BN037470.D	11 Jul 2025 12:59		RC/JU	Ok
7	SSTDICC1.6	SSTDICC1.6	BN037471.D	11 Jul 2025 13:36		RC/JU	Ok
8	SSTDICC3.2	SSTDICC3.2	BN037472.D	11 Jul 2025 14:12		RC/JU	Ok
9	SSTDICC5.0	SSTDICC5.0	BN037473.D	11 Jul 2025 14:49	Compound#20,23 removed from 5 ppm	RC/JU	Ok
10	SSTDICV0.4	ICVBN071125	BN037474.D	11 Jul 2025 15:25		RC/JU	Ok
11	PB168720BL	PB168720BL	BN037475.D	11 Jul 2025 16:02		RC/JU	Ok
12	Q2126-09	MDL-WATER-03-QT2-2	BN037476.D	11 Jul 2025 16:39	MDL-WATER -0.1 ppm	RC/JU	Ok,M
13	Q2126-09	MDL-WATER-03-QT2-2	BN037477.D	11 Jul 2025 17:16	MDL-WATER -0.2 ppm	RC/JU	Ok
14	Q2126-03	MDL-SOIL-03-QT2-202	BN037478.D	11 Jul 2025 17:52	MDL-SOIL -0.1 ppm	RC/JU	Ok
15	Q2126-03	MDL-SOIL-03-QT2-202	BN037479.D	11 Jul 2025 18:29	MDL-SOIL-0.2 PPM	RC/JU	Ok
16	PB168721BL	PB168721BL	BN037480.D	11 Jul 2025 19:06		RC/JU	Ok
17	SSTDCCC0.4	SSTDCCC0.4EC	BN037481.D	11 Jul 2025 19:43		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN071425

Review By	Rahul	Review On	7/15/2025 1:04:59 PM
Supervise By	Jagrut	Supervise On	7/15/2025 2:43:18 PM
SubDirectory	BN071425	HP Acquire Method	BNA_N, 8270_HP Processing Method bn071125

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6842, SP6843, SP6844, SP6845, SP6846, SP6847, SP6848
CCC	SP6846
Internal Standard/PEM	SP6740, 1ul/100ul sample
ICV/I.BLK	SP6854
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN037482.D	14 Jul 2025 08:31		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN037483.D	14 Jul 2025 09:10		RC/JU	Ok
3	PB168696BL	PB168696BL	BN037484.D	14 Jul 2025 09:49		RC/JU	Ok
4	Q2488-01	RW8-SP100-20250701	BN037485.D	14 Jul 2025 10:26		RC/JU	Ok
5	Q2488-02	RW8-SP303-20250701	BN037486.D	14 Jul 2025 11:02		RC/JU	Ok
6	PB168720BS	PB168720BS	BN037487.D	14 Jul 2025 11:39		RC/JU	Ok
7	PB168720BSD	PB168720BSD	BN037488.D	14 Jul 2025 12:15		RC/JU	Ok,M
8	Q2580-02	WC-URBAN-FILL-B2	BN037489.D	14 Jul 2025 13:27	Concentrated Matrix, Analyze with 10X dilution as a first run.	RC/JU	Not Ok
9	Q2580-02	WC-URBAN-FILL-B2	BN037490.D	14 Jul 2025 14:21	Need further 10X dilution	RC/JU	Dilution
10	Q2580-02MS	WC-URBAN-FILL-B2M	BN037491.D	14 Jul 2025 14:57	Internal Standard Failed	RC/JU	Ok,M
11	Q2580-02MSD	WC-URBAN-FILL-B2M	BN037492.D	14 Jul 2025 15:34	Internal Standard Failed	RC/JU	Ok,M
12	Q2580-02DL	WC-URBAN-FILL-B2DL	BN037493.D	14 Jul 2025 16:11		RC/JU	Ok,M
13	PB168831BS	PB168831BS	BN037494.D	14 Jul 2025 16:47		RC/JU	Ok,M
14	PB168831BL	PB168831BL	BN037495.D	14 Jul 2025 17:24		RC/JU	Ok
15	SSTDCCC0.4	SSTDCCC0.4EC	BN037496.D	14 Jul 2025 18:00		RC/JU	Ok

M : Manual Integration