

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2488

Client ID : SSTDICCC0.4

Date Analyzed: 07/11/2025

Lab File ID: BN037469.D

Time Analyzed: 12:22

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1754	7.732	4116	10.51	1977	14.37
UPPER LIMIT	3508	8.232	8232	11.009	3954	14.866
LOWER LIMIT	877	7.232	2058	10.009	988.5	13.866
EPA SAMPLE NO.						
01 PB168720BL	1870	7.73	4442	10.51	1998	14.37

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.