

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2488

Client ID : SSTDCCC0.4

Date Analyzed: 07/14/2025

Lab File ID: BN037483.D

Time Analyzed: 09:10

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1935	7.732	4597	10.51	2166	14.36
UPPER LIMIT	3870	8.232	9194	11.009	4332	14.855
LOWER LIMIT	967.5	7.232	2298.5	10.009	1083	13.855
EPA SAMPLE NO.						
01 PB168720BS	1966	7.73	4534	10.51	2005	14.36
02 PB168720BSD	1906	7.73	4315	10.51	1930	14.36
03 RW8-SP100-20250701	2365	7.73	5646	10.51	2637	14.36
04 RW8-SP303-20250701	1815	7.73	4225	10.51	1935	14.36

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.