

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2488

Client ID: SSTDICCC0.4

Date Analyzed: 07/11/2025

Lab File ID: BN037469.D

Time Analyzed: 12:22

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

01

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	3577	17.099	2684	21.286	2478	23.528
UPPER LIMIT	7154	17.599	5368	21.786	4956	24.028
LOWER LIMIT	1788.5	16.599	1342	20.786	1239	23.028
EPA SAMPLE NO.						
PB168720BL	3362	17.10	2365	21.29	2144	23.53

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.