

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN037465.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q2488
DFTPP Injection Date: 07/11/2025
DFTPP Injection Time: 08:36

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.2 (0.6) 1
69	Mass 69 relative abundance	39.7
70	Less than 2.0% of mass 69	0.1 (0.3) 1
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
365	Greater than 1% of mass 198	3.9
441	Present, but less than mass 443	11.9
442	Greater than 50% of mass 198	77.3
443	15.0 - 24.0% of mass 442	15 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN037467.D	07/11/2025	11:10
SSTDICC0.2	SSTDICC0.2	BN037468.D	07/11/2025	11:46
SSTDICCC0.4	SSTDICCC0.4	BN037469.D	07/11/2025	12:22
SSTDICC0.8	SSTDICC0.8	BN037470.D	07/11/2025	12:59
SSTDICC1.6	SSTDICC1.6	BN037471.D	07/11/2025	13:36
SSTDICC3.2	SSTDICC3.2	BN037472.D	07/11/2025	14:12
SSTDICC5.0	SSTDICC5.0	BN037473.D	07/11/2025	14:49
PB168720BL	PB168720BL	BN037475.D	07/11/2025	16:02
SSTDCCC0.4EC	SSTDCCC0.4	BN037481.D	07/11/2025	19:43