



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Walsh Construction Company II, LLC
Project: Construction of Shafts 17B-18B - PN 220084

SDG No.: Q2489
RunNo.: LB136375

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	7.01	7	100	90-110	07/03/2025
Sample ID: Corrosivity	CCV1	pH	2.01	2.00	101	90-110	07/03/2025
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	07/03/2025

Initial and Continuing Calibration Verification

Client: Walsh Construction Company II, LLC

SDG No.: Q2489

Project: Construction of Shafts 17B-18B - PN 220084

RunNo.: LB136389

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.099	0.099	100	85-115	07/07/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	07/07/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	07/07/2025
Sample ID: CCV3 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	07/07/2025