

Initial and Continuing Calibration Blank Summary

Client: Walsh Construction Company II, LLC
Project: Construction of Shafts 17B-18B - PN 220084

SDG No.: Q2489
RunNo.: LB136389

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Reactive Cyanide	mg/L	0.00097	0.0025	J	0.00096	0.005	07/07/2025
Sample ID: CCB1 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	07/07/2025
Sample ID: CCB2 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	07/07/2025
Sample ID: CCB3 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	07/07/2025