



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: Q2489

Contract: WALS01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB24	Mercury	0.076	+/-0.2	U	0.20	CV	07/11/2025	10:37	LB136435

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: Q2489

Contract: WALS01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB78	Mercury	0.076	+/-0.2	U	0.20	CV	07/11/2025	10:41	LB136435
CCB79	Mercury	0.076	+/-0.2	U	0.20	CV	07/11/2025	11:09	LB136435
CCB80	Mercury	0.076	+/-0.2	U	0.20	CV	07/11/2025	11:51	LB136435
CCB81	Mercury	0.076	+/-0.2	U	0.20	CV	07/11/2025	12:11	LB136435
CCB82	Mercury	0.076	+/-0.2	U	0.20	CV	07/11/2025	12:30	LB136435

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: Q2489

Contract: WALS01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	07/17/2025	00:46	LB136514
	Antimony	6.76	+/-25	U	50.0	P	07/17/2025	00:46	LB136514
	Arsenic	5.12	+/-10	U	20.0	P	07/17/2025	00:46	LB136514
	Barium	14.6	+/-50	U	100	P	07/17/2025	00:46	LB136514
	Beryllium	0.56	+/-3	U	6.00	P	07/17/2025	00:46	LB136514
	Boron	15.7	+/-50	U	100	P	07/17/2025	00:46	LB136514
	Cadmium	0.50	+/-3	U	6.00	P	07/17/2025	00:46	LB136514
	Calcium	234	+/-1000	U	2000	P	07/17/2025	00:46	LB136514
	Chromium	2.12	+/-5	U	10.0	P	07/17/2025	00:46	LB136514
	Cobalt	2.26	+/-15	U	30.0	P	07/17/2025	00:46	LB136514
	Copper	4.60	+/-10	U	20.0	P	07/17/2025	00:46	LB136514
	Iron	23.4	+/-50	U	100	P	07/17/2025	00:46	LB136514
	Lead	2.30	+/-6	U	12.0	P	07/17/2025	00:46	LB136514
	Magnesium	244	+/-1000	U	2000	P	07/17/2025	00:46	LB136514
	Manganese	5.94	+/-10	U	20.0	P	07/17/2025	00:46	LB136514
	Nickel	3.06	+/-20	U	40.0	P	07/17/2025	00:46	LB136514
	Potassium	918	+/-1000	U	2000	P	07/17/2025	00:46	LB136514
	Selenium	9.64	+/-10	U	20.0	P	07/17/2025	00:46	LB136514
	Silver	1.62	+/-5	U	10.0	P	07/17/2025	00:46	LB136514
	Sodium	868	+/-1000	U	2000	P	07/17/2025	00:46	LB136514
	Strontium	4.63	+/-10	U	20.0	P	07/17/2025	00:46	LB136514
	Lithium	15.6	+/-10	U	20.0	P	07/17/2025	00:46	LB136514
	Thallium	4.38	+/-20	U	40.0	P	07/17/2025	00:46	LB136514
	Tin	5.14	+/-20	U	40.0	P	07/17/2025	00:46	LB136514
	Titanium	6.42	+/-20	U	40.0	P	07/17/2025	00:46	LB136514
	Vanadium	6.26	+/-20	U	40.0	P	07/17/2025	00:46	LB136514
	Zinc	16.7	+/-20	U	40.0	P	07/17/2025	00:46	LB136514

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: Q2489

Contract: WALS01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	07/17/2025	01:16	LB136514
	Antimony	6.76	+/-25	U	50.0	P	07/17/2025	01:16	LB136514
	Arsenic	5.12	+/-10	U	20.0	P	07/17/2025	01:16	LB136514
	Barium	14.6	+/-50	U	100	P	07/17/2025	01:16	LB136514
	Beryllium	0.56	+/-3	U	6.00	P	07/17/2025	01:16	LB136514
	Boron	15.7	+/-50	U	100	P	07/17/2025	01:16	LB136514
	Cadmium	0.50	+/-3	U	6.00	P	07/17/2025	01:16	LB136514
	Calcium	234	+/-1000	U	2000	P	07/17/2025	01:16	LB136514
	Chromium	2.12	+/-5	U	10.0	P	07/17/2025	01:16	LB136514
	Cobalt	2.26	+/-15	U	30.0	P	07/17/2025	01:16	LB136514
	Copper	4.60	+/-10	U	20.0	P	07/17/2025	01:16	LB136514
	Iron	23.4	+/-50	U	100	P	07/17/2025	01:16	LB136514
	Lead	2.30	+/-6	U	12.0	P	07/17/2025	01:16	LB136514
	Magnesium	244	+/-1000	U	2000	P	07/17/2025	01:16	LB136514
	Manganese	5.94	+/-10	U	20.0	P	07/17/2025	01:16	LB136514
	Nickel	3.06	+/-20	U	40.0	P	07/17/2025	01:16	LB136514
	Potassium	918	+/-1000	U	2000	P	07/17/2025	01:16	LB136514
	Selenium	9.64	+/-10	U	20.0	P	07/17/2025	01:16	LB136514
	Silver	1.62	+/-5	U	10.0	P	07/17/2025	01:16	LB136514
	Sodium	868	+/-1000	U	2000	P	07/17/2025	01:16	LB136514
	Strontium	4.63	+/-10	U	20.0	P	07/17/2025	01:16	LB136514
	Lithium	15.6	+/-10	U	20.0	P	07/17/2025	01:16	LB136514
	Thallium	4.38	+/-20	U	40.0	P	07/17/2025	01:16	LB136514
	Tin	5.14	+/-20	U	40.0	P	07/17/2025	01:16	LB136514
	Titanium	6.42	+/-20	U	40.0	P	07/17/2025	01:16	LB136514
	Vanadium	6.26	+/-20	U	40.0	P	07/17/2025	01:16	LB136514
	Zinc	16.7	+/-20	U	40.0	P	07/17/2025	01:16	LB136514
CCB02	Aluminum	11.3	+/-50	U	100	P	07/17/2025	02:09	LB136514
	Antimony	6.76	+/-25	U	50.0	P	07/17/2025	02:09	LB136514
	Arsenic	5.12	+/-10	U	20.0	P	07/17/2025	02:09	LB136514
	Barium	14.6	+/-50	U	100	P	07/17/2025	02:09	LB136514
	Beryllium	0.56	+/-3	U	6.00	P	07/17/2025	02:09	LB136514
	Boron	15.7	+/-50	U	100	P	07/17/2025	02:09	LB136514
	Cadmium	0.50	+/-3	U	6.00	P	07/17/2025	02:09	LB136514
	Calcium	234	+/-1000	U	2000	P	07/17/2025	02:09	LB136514
	Chromium	2.12	+/-5	U	10.0	P	07/17/2025	02:09	LB136514
	Cobalt	2.26	+/-15	U	30.0	P	07/17/2025	02:09	LB136514
	Copper	4.60	+/-10	U	20.0	P	07/17/2025	02:09	LB136514
	Iron	23.4	+/-50	U	100	P	07/17/2025	02:09	LB136514

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: Q2489

Contract: WALS01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Lead	2.30	+/-6	U	12.0	P	07/17/2025	02:09	LB136514
	Magnesium	244	+/-1000	U	2000	P	07/17/2025	02:09	LB136514
	Manganese	5.94	+/-10	U	20.0	P	07/17/2025	02:09	LB136514
	Nickel	3.06	+/-20	U	40.0	P	07/17/2025	02:09	LB136514
	Potassium	918	+/-1000	U	2000	P	07/17/2025	02:09	LB136514
	Selenium	9.64	+/-10	U	20.0	P	07/17/2025	02:09	LB136514
	Silver	1.62	+/-5	U	10.0	P	07/17/2025	02:09	LB136514
	Sodium	868	+/-1000	U	2000	P	07/17/2025	02:09	LB136514
	Strontium	4.63	+/-10	U	20.0	P	07/17/2025	02:09	LB136514
	Lithium	15.6	+/-10	U	20.0	P	07/17/2025	02:09	LB136514
	Thallium	4.38	+/-20	U	40.0	P	07/17/2025	02:09	LB136514
	Tin	5.14	+/-20	U	40.0	P	07/17/2025	02:09	LB136514
	Titanium	6.42	+/-20	U	40.0	P	07/17/2025	02:09	LB136514
	Vanadium	6.26	+/-20	U	40.0	P	07/17/2025	02:09	LB136514
	Zinc	16.7	+/-20	U	40.0	P	07/17/2025	02:09	LB136514
CCB03	Aluminum	11.3	+/-50	U	100	P	07/17/2025	02:50	LB136514
	Antimony	6.76	+/-25	U	50.0	P	07/17/2025	02:50	LB136514
	Arsenic	5.12	+/-10	U	20.0	P	07/17/2025	02:50	LB136514
	Barium	14.6	+/-50	U	100	P	07/17/2025	02:50	LB136514
	Beryllium	0.56	+/-3	U	6.00	P	07/17/2025	02:50	LB136514
	Boron	15.7	+/-50	U	100	P	07/17/2025	02:50	LB136514
	Cadmium	0.50	+/-3	U	6.00	P	07/17/2025	02:50	LB136514
	Calcium	234	+/-1000	U	2000	P	07/17/2025	02:50	LB136514
	Chromium	2.12	+/-5	U	10.0	P	07/17/2025	02:50	LB136514
	Cobalt	2.26	+/-15	U	30.0	P	07/17/2025	02:50	LB136514
	Copper	4.60	+/-10	U	20.0	P	07/17/2025	02:50	LB136514
	Iron	23.4	+/-50	U	100	P	07/17/2025	02:50	LB136514
	Lead	2.30	+/-6	U	12.0	P	07/17/2025	02:50	LB136514
	Magnesium	244	+/-1000	U	2000	P	07/17/2025	02:50	LB136514
	Manganese	5.94	+/-10	U	20.0	P	07/17/2025	02:50	LB136514
	Nickel	3.06	+/-20	U	40.0	P	07/17/2025	02:50	LB136514
	Potassium	918	+/-1000	U	2000	P	07/17/2025	02:50	LB136514
	Selenium	9.64	+/-10	U	20.0	P	07/17/2025	02:50	LB136514
	Silver	1.62	+/-5	U	10.0	P	07/17/2025	02:50	LB136514
	Sodium	868	+/-1000	U	2000	P	07/17/2025	02:50	LB136514
	Strontium	4.63	+/-10	U	20.0	P	07/17/2025	02:50	LB136514
	Lithium	15.6	+/-10	U	20.0	P	07/17/2025	02:50	LB136514
	Thallium	4.38	+/-20	U	40.0	P	07/17/2025	02:50	LB136514
	Tin	5.14	+/-20	U	40.0	P	07/17/2025	02:50	LB136514

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: Q2489

Contract: WALS01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB03	Titanium	6.42	+/-20	U	40.0	P	07/17/2025	02:50	LB136514
	Vanadium	6.26	+/-20	U	40.0	P	07/17/2025	02:50	LB136514
	Zinc	16.7	+/-20	U	40.0	P	07/17/2025	02:50	LB136514