

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Walsh Construction Company II, LLC

SDG No.: Q2489

Contract: WALS01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	260000	255000	102	216000	294000	07/17/2025	00:55	LB136514
	Antimony	-1.45			-50	50	07/17/2025	00:55	LB136514
	Arsenic	0.92			-20	20	07/17/2025	00:55	LB136514
	Barium	3.84	6.0	64	-94	106	07/17/2025	00:55	LB136514
	Beryllium	1.56			-6	6	07/17/2025	00:55	LB136514
	Boron	8.12	1000	1	-100	100	07/17/2025	00:55	LB136514
	Cadmium	-1.26	1.0	126	-5	7	07/17/2025	00:55	LB136514
	Calcium	251000	245000	102	208000	282000	07/17/2025	00:55	LB136514
	Chromium	43.0	52.0	83	42	62	07/17/2025	00:55	LB136514
	Cobalt	1.97			-30	30	07/17/2025	00:55	LB136514
	Copper	0.57	2.0	28	-18	22	07/17/2025	00:55	LB136514
	Iron	102000	101000	101	85600	116500	07/17/2025	00:55	LB136514
	Lead	-4.61			-12	12	07/17/2025	00:55	LB136514
	Magnesium	257000	255000	101	216000	294000	07/17/2025	00:55	LB136514
	Manganese	9.51	7.0	136	-13	27	07/17/2025	00:55	LB136514
	Nickel	2.34	2.0	117	-38	42	07/17/2025	00:55	LB136514
	Potassium	35.9			0	0	07/17/2025	00:55	LB136514
	Selenium	1.77			-20	20	07/17/2025	00:55	LB136514
	Silver	1.50			-10	10	07/17/2025	00:55	LB136514
	Sodium	-302			0	0	07/17/2025	00:55	LB136514
	Strontium	6.50			0	0	07/17/2025	00:55	LB136514
	Lithium	5.20	1000	1	-20	20	07/17/2025	00:55	LB136514
	Thallium	-9.09			-40	40	07/17/2025	00:55	LB136514
	Tin	-1.93	1000		-40	40	07/17/2025	00:55	LB136514
	Titanium	-5.55			0	0	07/17/2025	00:55	LB136514
	Vanadium	0.19			-40	40	07/17/2025	00:55	LB136514
	Zinc	1.98			-40	40	07/17/2025	00:55	LB136514
ICSAB01	Aluminum	253000	247000	102	209000	285000	07/17/2025	00:59	LB136514
	Antimony	623	618	101	525	711	07/17/2025	00:59	LB136514
	Arsenic	103	104	99	88.4	120	07/17/2025	00:59	LB136514
	Barium	520	537	97	437	637	07/17/2025	00:59	LB136514
	Beryllium	498	495	101	420	570	07/17/2025	00:59	LB136514
	Boron	965	1000	96	850	1150	07/17/2025	00:59	LB136514
	Cadmium	984	972	101	826	1120	07/17/2025	00:59	LB136514
	Calcium	247000	235000	105	199000	271000	07/17/2025	00:59	LB136514
	Chromium	515	542	95	460	624	07/17/2025	00:59	LB136514
	Cobalt	506	476	106	404	548	07/17/2025	00:59	LB136514
	Copper	498	511	98	434	588	07/17/2025	00:59	LB136514
	Iron	102000	99300	103	84400	114500	07/17/2025	00:59	LB136514
	Lead	41.8	49.0	85	37	61	07/17/2025	00:59	LB136514
	Magnesium	253000	248000	102	210000	286000	07/17/2025	00:59	LB136514
	Manganese	511	507	101	430	584	07/17/2025	00:59	LB136514

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Walsh Construction Company II, LLC

SDG No.: Q2489

Contract: WALS01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Nickel	998	954	105	810	1100	07/17/2025	00:59	LB136514
	Potassium	75.8			0	0	07/17/2025	00:59	LB136514
	Selenium	50.6	46.0	110	26	66	07/17/2025	00:59	LB136514
	Silver	221	201	110	170	232	07/17/2025	00:59	LB136514
	Sodium	-259			0	0	07/17/2025	00:59	LB136514
	Strontium	1060			0	0	07/17/2025	00:59	LB136514
	Lithium	1120	1000	112	850	1150	07/17/2025	00:59	LB136514
	Thallium	92.6	108	86	68	148	07/17/2025	00:59	LB136514
	Tin	1000	1000	100	850	1150	07/17/2025	00:59	LB136514
	Titanium	1030			0	0	07/17/2025	00:59	LB136514
	Vanadium	485	491	99	417	565	07/17/2025	00:59	LB136514
	Zinc	1080	952	113	809	1095	07/17/2025	00:59	LB136514