

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: Q2489

Contract: WALS01

Lab Code: ACE

Matrix: Water

Level: LOW

Client ID: G4(0-6)A

Sample ID: Q2489-01 **Spiked ID:** Q2489-01A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Potassium	ug/L	75 - 125	11700		5690		5000	120		P
Silver	ug/L	75 - 125	26.3		5.00	U	37.5	70	N	P