

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Walsh Construction Company II, LLC **Level:** LOW **SDG No.:** Q2489
Contract: WALS01 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q2489-01 **Client ID:** G4(0-6)DUP
Percent Solids for Sample: NA **Duplicate ID** Q2489-01DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	145		147		1		P
Antimony	ug/L	20	7.67	J	9.50	J	21		P
Arsenic	ug/L	20	10.0	U	10.0	U			P
Barium	ug/L	20	65.4		65.6		0		P
Beryllium	ug/L	20	3.00	U	3.00	U			P
Boron	ug/L	20	74.2		75.0		1		P
Cadmium	ug/L	20	1.27	J	1.25	J	2		P
Calcium	ug/L	20	610000		608000		0		P
Chromium	ug/L	20	2.80	J	3.15	J	12		P
Cobalt	ug/L	20	1.28	J	1.28	J	0		P
Copper	ug/L	20	19.3		18.9		2		P
Iron	ug/L	20	27.8	J	29.1	J	5		P
Lead	ug/L	20	8.88		9.55		7		P
Magnesium	ug/L	20	13000		13100		1		P
Manganese	ug/L	20	369		371		1		P
Mercury	ug/L	20	0.20	U	0.20	U			CV
Nickel	ug/L	20	8.31	J	8.27	J	0		P
Potassium	ug/L	20	5690		5610		1		P
Selenium	ug/L	20	10.0	U	10.0	U			P
Silver	ug/L	20	5.00	U	5.00	U			P
Sodium	ug/L	20	20600		20300		1		P
Strontium	ug/L	20	2580		2560		1		P
Lithium	ug/L	20	74.3		75.5		2		P
Thallium	ug/L	20	20.0	U	20.0	U			P
Tin	ug/L	20	20.0	U	20.0	U			P
Titanium	ug/L	20	20.0	U	20.0	U			P
Vanadium	ug/L	20	20.0	U	20.0	U			P
Zinc	ug/L	20	225		221		2		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Walsh Construction Company II, LLC **Level:** LOW **SDG No.:** Q2489
Contract: WALS01 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q2489-01MS **Client ID:** G4(0-6)MSD
Percent Solids for Sample: NA **Duplicate ID** Q2489-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	1140		1130		1		P
Antimony	ug/L	20	456		455		0		P
Arsenic	ug/L	20	435		435		0		P
Barium	ug/L	20	172		171		1		P
Beryllium	ug/L	20	92.4		91.8		1		P
Boron	ug/L	20	205		203		1		P
Cadmium	ug/L	20	95.5		96.0		1		P
Calcium	ug/L	20	634000		632000		0		P
Chromium	ug/L	20	182		181		1		P
Cobalt	ug/L	20	99.2		99.5		0		P
Copper	ug/L	20	167		167		0		P
Iron	ug/L	20	1640		1630		1		P
Lead	ug/L	20	460		462		0		P
Magnesium	ug/L	20	14400		14400		0		P
Manganese	ug/L	20	482		481		0		P
Mercury	ug/L	20	4.20		4.29		2		CV
Nickel	ug/L	20	249		250		0		P
Potassium	ug/L	20	12100		12000		1		P
Selenium	ug/L	20	1130		1130		0		P
Silver	ug/L	20	26.8		25.0		7		P
Sodium	ug/L	20	23900		23500		2		P
Strontium	ug/L	20	2770		2740		1		P
Lithium	ug/L	20	187		182		3		P
Thallium	ug/L	20	872		859		2		P
Tin	ug/L	20	338		341		1		P
Titanium	ug/L	20	76.6		77.2		1		P
Vanadium	ug/L	20	148		148		0		P
Zinc	ug/L	20	345		341		1		P