

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: Q2489

Contract: WALS01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168773BS							
Aluminum	ug/L	1000	999		100	80 - 120	P
Antimony	ug/L	400	416		104	80 - 120	P
Arsenic	ug/L	400	394		98	80 - 120	P
Barium	ug/L	100	103		103	80 - 120	P
Beryllium	ug/L	100	102		102	80 - 120	P
Boron	ug/L	150	152		101	80 - 120	P
Cadmium	ug/L	100	94.7		95	80 - 120	P
Calcium	ug/L	500	549	J	110	80 - 120	P
Chromium	ug/L	200	189		94	80 - 120	P
Cobalt	ug/L	100	97.8		98	80 - 120	P
Copper	ug/L	150	155		103	80 - 120	P
Iron	ug/L	1500	1500		100	80 - 120	P
Lead	ug/L	500	475		95	80 - 120	P
Magnesium	ug/L	1000	1000		100	80 - 120	P
Manganese	ug/L	100	103		103	80 - 120	P
Nickel	ug/L	250	245		98	80 - 120	P
Potassium	ug/L	5000	5060		101	80 - 120	P
Selenium	ug/L	1000	992		99	80 - 120	P
Silver	ug/L	37.5	38.0		101	80 - 120	P
Sodium	ug/L	1500	1580		105	80 - 120	P
Strontium	ug/L	100	100		100	80 - 120	P
Lithium	ug/L	100	96.7		97	80 - 120	P
Thallium	ug/L	1000	983		98	80 - 120	P
Tin	ug/L	350	343		98	80 - 120	P
Titanium	ug/L	100	103		103	80 - 120	P
Vanadium	ug/L	150	146		97	80 - 120	P
Zinc	ug/L	100	107		107	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: Q2489

Contract: WALS01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168814BS Mercury	ug/L	4.0	4.70		118	80 - 120	CV