

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136403

Review By	MOHAN	Review On	7/10/2025 9:48:52 AM
Supervise By	jaswal	Supervise On	7/10/2025 11:31:40 PM
STD. NAME	STD REF.#		
ICAL Standard	MP86306,MP86307,MP86308,MP86309,MP86310,MP86311		
ICV Standard	MP86312		
CCV Standard	MP86314		
ICSA Standard			
CRI Standard	MP86316		
LCS Standard			
Chk Standard	MP86313,MP86315,MP86317,MP86335		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	07/09/25 11:26		MOHAN	OK
2	S0.2	S0.2	CAL2	07/09/25 11:28		MOHAN	OK
3	S2.5	S2.5	CAL3	07/09/25 11:31		MOHAN	OK
4	S5	S5	CAL4	07/09/25 11:33		MOHAN	OK
5	S7.5	S7.5	CAL5	07/09/25 11:36		MOHAN	OK
6	S10	S10	CAL6	07/09/25 11:38		MOHAN	OK
7	ICV22	ICV22	ICV	07/09/25 11:43		MOHAN	OK
8	ICB22	ICB22	ICB	07/09/25 11:46		MOHAN	OK
9	CCV70	CCV70	CCV	07/09/25 11:48		MOHAN	OK
10	CCB70	CCB70	CCB	07/09/25 11:50		MOHAN	OK
11	CRA	CRA	CRDL	07/09/25 11:53		MOHAN	OK
12	HighStd	HighStd	HIGH STD	07/09/25 11:55		MOHAN	OK
13	ChkStd	ChkStd	SAM	07/09/25 11:57		MOHAN	OK
14	PB168760BL	PB168760BL	MB	07/09/25 12:00		MOHAN	OK
15	PB168760BS	PB168760BS	LCS	07/09/25 12:02		MOHAN	OK
16	Q2489-01	G4(0-6)	SAM	07/09/25 12:04		MOHAN	OK
17	Q2489-02	G4(6-12)	SAM	07/09/25 12:06		MOHAN	OK
18	Q2489-03	G3(0-6)	SAM	07/09/25 12:09		MOHAN	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136403

Review By	MOHAN	Review On	7/10/2025 9:48:52 AM
Supervise By	jaswal	Supervise On	7/10/2025 11:31:40 PM
STD. NAME	STD REF.#		
ICAL Standard	MP86306,MP86307,MP86308,MP86309,MP86310,MP86311		
ICV Standard	MP86312		
CCV Standard	MP86314		
ICSA Standard			
CRI Standard	MP86316		
LCS Standard			
Chk Standard	MP86313,MP86315,MP86317,MP86335		

19	Q2489-04	G3(6-12)	SAM	07/09/25 12:14		MOHAN	OK
20	Q2489-05	G2(0-6)	SAM	07/09/25 12:16		MOHAN	OK
21	CCV71	CCV71	CCV	07/09/25 12:18		MOHAN	OK
22	CCB71	CCB71	CCB	07/09/25 12:20		MOHAN	OK
23	Q2489-06	G2(6-12)	SAM	07/09/25 12:23		MOHAN	OK
24	Q2489-07	G1(0-6)	SAM	07/09/25 12:25		MOHAN	OK
25	Q2489-08	G1(6-12)	SAM	07/09/25 12:27		MOHAN	OK
26	Q2489-08DUP	G1(6-12)DUP	DUP	07/09/25 12:30		MOHAN	OK
27	Q2489-08MS	G1(6-12)MS	MS	07/09/25 12:32		MOHAN	OK
28	Q2489-08MSD	G1(6-12)MSD	MSD	07/09/25 12:34		MOHAN	OK
29	PB168766BL	PB168766BL	MB	07/09/25 12:37		MOHAN	OK
30	PB168766BS	PB168766BS	LCS	07/09/25 12:39		MOHAN	OK
31	Q2503-01	GCAP2W	SAM	07/09/25 12:41	Hg is high	MOHAN	Dilution
32	Q2509-01	AUD-25-0110-0111	SAM	07/09/25 12:48		MOHAN	OK
33	CCV72	CCV72	CCV	07/09/25 12:56		MOHAN	OK
34	CCB72	CCB72	CCB	07/09/25 12:58		MOHAN	OK
35	Q2509-01DUP	AUD-25-0110-0111DU	DUP	07/09/25 13:01		MOHAN	OK
36	Q2509-01MS	AUD-25-0110-0111MS	MS	07/09/25 13:07		MOHAN	OK
37	Q2509-01MSD	AUD-25-0110-0111MS	MSD	07/09/25 13:09		MOHAN	OK
38	PB168728TB	PB168728TB	MB	07/09/25 13:12		MOHAN	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136403

Review By	MOHAN	Review On	7/10/2025 9:48:52 AM
Supervise By	jaswal	Supervise On	7/10/2025 11:31:40 PM
STD. NAME	STD REF.#		
ICAL Standard	MP86306,MP86307,MP86308,MP86309,MP86310,MP86311		
ICV Standard	MP86312		
CCV Standard	MP86314		
ICSA Standard			
CRI Standard	MP86316		
LCS Standard			
Chk Standard	MP86313,MP86315,MP86317,MP86335		

39	Q2489-08L	G1(6-12)L	SD	07/09/25 13:14		MOHAN	OK
40	Q2489-08A	G1(6-12)A	PS	07/09/25 13:16		MOHAN	OK
41	Q2509-01L	AUD-25-0110-0111L	SD	07/09/25 13:18		MOHAN	OK
42	Q2509-01A	AUD-25-0110-0111A	PS	07/09/25 13:21		MOHAN	OK
43	Q2503-01DL	GCAP2WDL	SAM	07/09/25 13:28	25X for Hg	MOHAN	Confirms
44	CCV73	CCV73	CCV	07/09/25 13:31		MOHAN	OK
45	CCB73	CCB73	CCB	07/09/25 13:33		MOHAN	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136511

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	07/16/25 12:55		Jaswal	OK
2	S1	S1	CAL2	07/16/25 13:00		Jaswal	OK
3	S2	S2	CAL3	07/16/25 13:04		Jaswal	OK
4	S3	S3	CAL4	07/16/25 13:08		Jaswal	OK
5	S4	S4	CAL5	07/16/25 13:12		Jaswal	OK
6	S5	S5	CAL6	07/16/25 13:16		Jaswal	OK
7	ICV01	ICV01	ICV	07/16/25 13:21	ICV fail for Pb,Na (200.7) (95-105)	Jaswal	OK
8	LLICV01	LLICV01	LLICV	07/16/25 13:53		Jaswal	OK
9	ICB01	ICB01	ICB	07/16/25 14:11		Jaswal	OK
10	CRI01	CRI01	CRDL	07/16/25 14:22		Jaswal	OK
11	ICSA01	ICSA01	ICSA	07/16/25 14:26		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	07/16/25 14:38		Jaswal	OK
13	ICSADL	ICSADL	ICSA	07/16/25 14:43		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	07/16/25 14:47		Jaswal	OK
15	CCV01	CCV01	CCV	07/16/25 14:53	CCV fail for Na	Jaswal	OK
16	CCB01	CCB01	CCB	07/16/25 14:57		Jaswal	OK
17	PB168761TB	PB168761TB	MB	07/16/25 15:01		Jaswal	OK
18	PB168728TB	PB168728TB	MB	07/16/25 15:08		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136511

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

19	Q2489-08	G1(6-12)	SAM	07/16/25 15:13		Jaswal	OK
20	Q2489-08DUP	G1(6-12)DUP	DUP	07/16/25 15:17		Jaswal	OK
21	Q2489-08L	G1(6-12)L	SD	07/16/25 15:22		Jaswal	OK
22	Q2489-08MS	G1(6-12)MS	MS	07/16/25 15:26		Jaswal	OK
23	Q2489-08MSD	G1(6-12)MSD	MSD	07/16/25 15:30		Jaswal	OK
24	Q2489-08A	G1(6-12)A	PS	07/16/25 15:34		Jaswal	OK
25	CCV02	CCV02	CCV	07/16/25 15:39	CCV fail for Na	Jaswal	OK
26	CCB02	CCB02	CCB	07/16/25 15:43		Jaswal	OK
27	Q2593-02	002 35th AVE (MAY)	SAM	07/16/25 15:48		Jaswal	OK
28	Q2593-02DUP	002 35th AVE (MAY)D	DUP	07/16/25 15:57		Jaswal	OK
29	Q2593-02L	002 35th AVE (MAY)L	SD	07/16/25 16:01		Jaswal	OK
30	Q2593-02MS	002 35th AVE (MAY)M	MS	07/16/25 16:05		Jaswal	OK
31	Q2593-02MSD	002 35th AVE (MAY)M	MSD	07/16/25 16:09		Jaswal	OK
32	Q2593-02A	002 35th AVE (MAY)A	PS	07/16/25 16:13		Jaswal	OK
33	PB168843BL	PB168843BL	MB	07/16/25 16:17		Jaswal	OK
34	PB168843BS	PB168843BS	LCS	07/16/25 16:21		Jaswal	OK
35	CCV03	CCV03	CCV	07/16/25 16:25	CCV fail for Na	Jaswal	OK
36	CCB03	CCB03	CCB	07/16/25 16:30		Jaswal	OK
37	Q2578-03	WC-A5-01-C	SAM	07/16/25 16:46		Jaswal	OK
38	Q2578-07	WC-A2-09-C	SAM	07/16/25 16:50		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136511

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

39	Q2578-11	WC-A2-10-C	SAM	07/16/25 16:54		Jaswal	OK
40	Q2578-15	WC-A2-11-C	SAM	07/16/25 16:59		Jaswal	OK
41	Q2578-19	WC-A2-12-C	SAM	07/16/25 17:05		Jaswal	OK
42	Q2579-03	WC-A2-13-C	SAM	07/16/25 17:10		Jaswal	OK
43	Q2579-07	WC-A2-14-C	SAM	07/16/25 17:14		Jaswal	OK
44	PB168803TB	PB168803TB	MB	07/16/25 17:18		Jaswal	OK
45	Q2561-07	AUD-25-0117	SAM	07/16/25 17:23		Jaswal	OK
46	Q2565-06	MOO-25-0194-0195	SAM	07/16/25 17:27		Jaswal	OK
47	CCV04	CCV04	CCV	07/16/25 17:32	CCV fail for Na	Jaswal	OK
48	CCB04	CCB04	CCB	07/16/25 17:40		Jaswal	OK
49	Q2565-07	MOO-25-0191	SAM	07/16/25 17:44		Jaswal	OK
50	Q2565-08	MOO-25-0196	SAM	07/16/25 17:49		Jaswal	OK
51	Q2565-09	MOO-25-0180	SAM	07/16/25 17:53		Jaswal	OK
52	Q2571-04	TP-18	SAM	07/16/25 17:58		Jaswal	OK
53	Q2571-08	TP-17	SAM	07/16/25 18:02		Jaswal	OK
54	Q2586-04	TP-16	SAM	07/16/25 18:06		Jaswal	OK
55	Q2586-04DUP	TP-16DUP	DUP	07/16/25 18:11		Jaswal	OK
56	Q2586-04L	TP-16L	SD	07/16/25 18:15		Jaswal	OK
57	Q2586-04MS	TP-16MS	MS	07/16/25 18:20		Jaswal	OK
58	Q2586-04MSD	TP-16MSD	MSD	07/16/25 18:24		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136511

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

59	CCV05	CCV05	CCV	07/16/25 18:28	CCV fail for Na	Jaswal	OK
60	CCB05	CCB05	CCB	07/16/25 18:35		Jaswal	OK
61	Q2586-04A	TP-16A	PS	07/16/25 18:39		Jaswal	OK
62	PB168857BL	PB168857BL	MB	07/16/25 18:43		Jaswal	OK
63	PB168857BS	PB168857BS	LCS	07/16/25 18:47		Jaswal	OK
64	CCV06	CCV06	CCV	07/16/25 18:51	CCV fail for Na	Jaswal	OK
65	CCB06	CCB06	CCB	07/16/25 18:56		Jaswal	OK



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID:

Daily Analysis Runlog For Sequence/QC Batch ID #

Review By		Review On					
Supervise By		Supervise On					
STD. NAME		STD REF.#					
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard							
Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status