



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Walsh Construction Company II, LLC

SDG No.: Q2489

Contract: WALS01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.18	0.2	90	70 - 130	CV	07/09/2025	11:53	LB136403
CRI01	Arsenic	22.0	20.0	110	65 - 135	P	07/10/2025	18:09	LB136519
	Barium	101	100	101	65 - 135	P	07/10/2025	18:09	LB136519
	Cadmium	5.94	6.0	99	65 - 135	P	07/10/2025	18:09	LB136519
	Chromium	10.7	10.0	106	65 - 135	P	07/10/2025	18:09	LB136519
	Lead	11.2	12.0	93	65 - 135	P	07/10/2025	18:09	LB136519
	Selenium	24.0	20.0	120	65 - 135	P	07/10/2025	18:09	LB136519
	Silver	8.69	10.0	87	65 - 135	P	07/10/2025	18:09	LB136519
CRI01	Arsenic	20.1	20.0	101	65 - 135	P	07/16/2025	14:22	LB136511
	Barium	100	100	100	65 - 135	P	07/16/2025	14:22	LB136511
	Cadmium	5.84	6.0	97	65 - 135	P	07/16/2025	14:22	LB136511
	Chromium	9.85	10.0	98	65 - 135	P	07/16/2025	14:22	LB136511
	Lead	11.0	12.0	92	65 - 135	P	07/16/2025	14:22	LB136511
	Selenium	18.3	20.0	92	65 - 135	P	07/16/2025	14:22	LB136511
	Silver	10.4	10.0	104	65 - 135	P	07/16/2025	14:22	LB136511