

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: Q2489

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168728TB		WATER		Batch Number:	PB168760		Prep Date:	07/08/2025	
	Mercury	0.76	<2	U	2.00	CV	07/09/2025	13:12	LB136403
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168760BL		WATER		Batch Number:	PB168760		Prep Date:	07/08/2025	
	Mercury	0.076	<0.2	U	0.20	CV	07/09/2025	12:00	LB136403

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: Q2489

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168728TB		WATER		Batch Number:	PB168744		Prep Date:	07/07/2025	
	Arsenic	25.6	<50	U	100	P	07/16/2025	15:08	LB136511
	Barium	72.8	<250	U	500	P	07/16/2025	15:08	LB136511
	Cadmium	2.50	<15	U	30.0	P	07/16/2025	15:08	LB136511
	Chromium	10.6	<25	U	50.0	P	07/16/2025	15:08	LB136511
	Lead	11.5	<30	U	60.0	P	07/16/2025	15:08	LB136511
	Selenium	48.2	<50	U	100	P	07/16/2025	15:08	LB136511
	Silver	8.10	<25	U	50.0	P	07/16/2025	15:08	LB136511
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168744BL		WATER		Batch Number:	PB168744		Prep Date:	07/07/2025	
	Arsenic	25.6	<50	U	100	P	07/10/2025	22:21	LB136519
	Barium	72.8	<250	U	500	P	07/10/2025	22:21	LB136519
	Cadmium	2.50	<15	U	30.0	P	07/10/2025	22:21	LB136519
	Chromium	10.6	<25	U	50.0	P	07/10/2025	22:21	LB136519
	Lead	11.5	<30	U	60.0	P	07/10/2025	22:21	LB136519
	Selenium	48.2	<50	U	100	P	07/10/2025	22:21	LB136519
	Silver	8.10	<25	U	50.0	P	07/10/2025	22:21	LB136519