

**Metals**  
**- 4 -**  
**INTERFERENCE CHECK SAMPLE**

**Client:** Walsh Construction Company II, LLC

**SDG No.:** Q2489

**Contract:** WALS01

**Lab Code:** ACE

**ICS Source:** EPA

**Instrument ID:** P4

| Sample ID      | Analyte  | Result<br>ug/L | True Value<br>ug/L | %<br>Recovery | Low<br>Limit<br>(ug/L) | High<br>Limit<br>(ug/L) | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|----------------|----------|----------------|--------------------|---------------|------------------------|-------------------------|------------------|------------------|---------------|
| <b>ICSA01</b>  | Arsenic  | 9.78           |                    |               | -20                    | 20                      | 07/10/2025       | 18:14            | LB136519      |
|                | Barium   | 1.97           | 6.0                | 33            | -94                    | 106                     | 07/10/2025       | 18:14            | LB136519      |
|                | Cadmium  | -0.15          | 1.0                | 15            | -5                     | 7                       | 07/10/2025       | 18:14            | LB136519      |
|                | Chromium | 58.9           | 52.0               | 113           | 42                     | 62                      | 07/10/2025       | 18:14            | LB136519      |
|                | Lead     | -4.02          |                    |               | -12                    | 12                      | 07/10/2025       | 18:14            | LB136519      |
|                | Selenium | -5.79          |                    |               | -20                    | 20                      | 07/10/2025       | 18:14            | LB136519      |
|                | Silver   | -0.013         |                    |               | -10                    | 10                      | 07/10/2025       | 18:14            | LB136519      |
| <b>ICSAB01</b> | Arsenic  | 105            | 104                | 101           | 88.4                   | 120                     | 07/10/2025       | 18:18            | LB136519      |
|                | Barium   | 487            | 537                | 91            | 437                    | 637                     | 07/10/2025       | 18:18            | LB136519      |
|                | Cadmium  | 952            | 972                | 98            | 826                    | 1120                    | 07/10/2025       | 18:18            | LB136519      |
|                | Chromium | 552            | 542                | 102           | 460                    | 624                     | 07/10/2025       | 18:18            | LB136519      |
|                | Lead     | 39.5           | 49.0               | 81            | 37                     | 61                      | 07/10/2025       | 18:18            | LB136519      |
|                | Selenium | 48.9           | 46.0               | 106           | 26                     | 66                      | 07/10/2025       | 18:18            | LB136519      |
|                | Silver   | 203            | 201                | 101           | 170                    | 232                     | 07/10/2025       | 18:18            | LB136519      |
| <b>ICSA01</b>  | Arsenic  | 9.98           |                    |               | -20                    | 20                      | 07/16/2025       | 14:26            | LB136511      |
|                | Barium   | 2.39           | 6.0                | 40            | -94                    | 106                     | 07/16/2025       | 14:26            | LB136511      |
|                | Cadmium  | -0.36          | 1.0                | 36            | -5                     | 7                       | 07/16/2025       | 14:26            | LB136511      |
|                | Chromium | 56.1           | 52.0               | 108           | 42                     | 62                      | 07/16/2025       | 14:26            | LB136511      |
|                | Lead     | -7.00          |                    |               | -12                    | 12                      | 07/16/2025       | 14:26            | LB136511      |
|                | Selenium | 9.85           |                    |               | -20                    | 20                      | 07/16/2025       | 14:26            | LB136511      |
|                | Silver   | 1.35           |                    |               | -10                    | 10                      | 07/16/2025       | 14:26            | LB136511      |
| <b>ICSAB01</b> | Arsenic  | 104            | 104                | 100           | 88.4                   | 120                     | 07/16/2025       | 14:38            | LB136511      |
|                | Barium   | 477            | 537                | 89            | 437                    | 637                     | 07/16/2025       | 14:38            | LB136511      |
|                | Cadmium  | 913            | 972                | 94            | 826                    | 1120                    | 07/16/2025       | 14:38            | LB136511      |
|                | Chromium | 483            | 542                | 89            | 460                    | 624                     | 07/16/2025       | 14:38            | LB136511      |
|                | Lead     | 38.4           | 49.0               | 78            | 37                     | 61                      | 07/16/2025       | 14:38            | LB136511      |
|                | Selenium | 60.4           | 46.0               | 131           | 26                     | 66                      | 07/16/2025       | 14:38            | LB136511      |
|                | Silver   | 204            | 201                | 102           | 170                    | 232                     | 07/16/2025       | 14:38            | LB136511      |