

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Walsh Construction Company II, LLC **Level:** LOW **SDG No.:** Q2489
Contract: WALS01 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q2489-08 **Client ID:** G1(6-12)DUP
Percent Solids for Sample: NA **Duplicate ID** Q2489-08DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	3170		3220		2		P
Cadmium	ug/L	20	30.0	U	30.0	U			P
Chromium	ug/L	20	50.0	U	50.0	U			P
Lead	ug/L	20	60.0	U	60.0	U			P
Mercury	ug/L	20	1.62	J	1.94	J	18		CV
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Walsh Construction Company II, LLC **Level:** LOW **SDG No.:** Q2489
Contract: WALS01 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q2489-08MS **Client ID:** G1(6-12)MSD
Percent Solids for Sample: NA **Duplicate ID** Q2489-08MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	3330		3400		2		P
Barium	ug/L	20	3470		3490		1		P
Cadmium	ug/L	20	778		791		2		P
Chromium	ug/L	20	1520		1560		3		P
Lead	ug/L	20	3700		3760		2		P
Mercury	ug/L	20	34.0		34.1		0		CV
Selenium	ug/L	20	8340		8560		3		P
Silver	ug/L	20	299		305		2		P