

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: Q2489

Contract: WALS01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168744BS							
Arsenic	ug/L	4000	4000		100	80 - 120	P
Barium	ug/L	1000	1010		101	80 - 120	P
Cadmium	ug/L	1000	982		98	80 - 120	P
Chromium	ug/L	2000	2100		105	80 - 120	P
Lead	ug/L	5000	4900		98	80 - 120	P
Selenium	ug/L	10000	10100		101	80 - 120	P
Silver	ug/L	380	360		95	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: Q2489

Contract: WALS01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168760BS Mercury	ug/L	4.0	3.61		90	80 - 120	CV