

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2489

Instrument ID: ECD_S

Calibration Date(s): 07/11/2025 07/11/2025

Calibration Times: 16:00 17:36

GC Column: RTX-CLP

ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS031006.D</u>	CF 500 =	<u>PS031007.D</u>		
CF 750 =		<u>PS031008.D</u>	CF 1000 =	<u>PS031009.D</u>	CF 1500 =	<u>PS031010.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	20737500000	18270800000	18189000000	18194100000	17662300000	18610700000	7
2,4-D	3556670000	2984170000	2931980000	2976850000	2942510000	3078440000	9
2,4-DCAA	4821770000	3910250000	3755950000	3730170000	3557220000	3955070000	13

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2489

Instrument ID: ECD_S

Calibration Date(s): 07/11/2025 07/11/2025

Calibration Times: 16:00 17:36

GC Column: RTX-CLP2

ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS031006.D</u>	CF 500 =	<u>PS031007.D</u>		
CF 750 =		<u>PS031008.D</u>	CF 1000 =	<u>PS031009.D</u>	CF 1500 =	<u>PS031010.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	17216300000	14510800000	14174600000	14131100000	13346200000	14675800000	10
2,4-D	2058060000	1665160000	1617550000	1610710000	1545850000	1699460000	12
2,4-DCAA	1258080000	1011280000	984071000	981302000	945431000	1036030000	12