

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2489
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/16/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031062.D **Time Analyzed:** 07:14

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.346	9.253	9.453	803.190	712.500	12.7
2,4-D	8.461	8.367	8.567	803.510	705.000	14.0
2,4-DCAA	7.327	7.233	7.433	616.170	750.000	-17.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2489
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/16/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031062.D **Time Analyzed:** 07:14

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.928	9.832	10.032	694.060	712.500	-2.6
2,4-D	9.024	8.927	9.127	764.900	705.000	8.5
2,4-DCAA	7.767	7.669	7.869	685.950	750.000	-8.5