

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2489
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/16/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031075.D **Time Analyzed:** 19:54

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.348	9.253	9.453	958.670	712.500	34.6
2,4-D	8.463	8.367	8.567	926.410	705.000	31.4
2,4-DCAA	7.328	7.233	7.433	921.480	750.000	22.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2489
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/16/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031075.D **Time Analyzed:** 19:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.932	9.832	10.032	796.010	712.500	11.7
2,4-D	9.027	8.927	9.127	776.390	705.000	10.1
2,4-DCAA	7.768	7.669	7.869	817.330	750.000	9.0