

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2489
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/17/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031086.D **Time Analyzed:** 00:19

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.347	9.253	9.453	986.520	712.500	38.5
2,4-D	8.463	8.367	8.567	901.110	705.000	27.8
2,4-DCAA	7.329	7.233	7.433	941.730	750.000	25.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2489
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/17/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031086.D **Time Analyzed:** 00:19

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.931	9.832	10.032	814.250	712.500	14.3
2,4-D	9.026	8.927	9.127	799.670	705.000	13.4
2,4-DCAA	7.768	7.669	7.869	835.520	750.000	11.4