

Surrogate Summary

SDG No.: Q2489

Client: Walsh Construction Company II, LLC

Analytical Method: 8151A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PS031005.D	PIBLK-PS031005.D	2,4-DCAA	1	500	356	71		61	136
		2,4-DCAA	2	500	496	99		61	136
I.BLK-PS031046.D	PIBLK-PS031046.D	2,4-DCAA	1	500	444	89		61	136
		2,4-DCAA	2	500	486	97		61	136
PB168825BL	PB168825BL	2,4-DCAA	1	500	366	73		61	136
		2,4-DCAA	2	500	442	88		61	136
PB168825BS	PB168825BS	2,4-DCAA	1	500	436	87		61	136
		2,4-DCAA	2	500	441	88		61	136
PB168825TB	PB168825TB	2,4-DCAA	1	500	364	73		61	136
		2,4-DCAA	2	500	454	91		61	136
Q2489-01	G4(0-6)	2,4-DCAA	1	500	416	83		61	136
		2,4-DCAA	2	500	510	102		61	136
Q2489-01MS	G4(0-6)MS	2,4-DCAA	1	500	411	82		61	136
		2,4-DCAA	2	500	383	77		61	136
Q2489-01MSD	G4(0-6)MSD	2,4-DCAA	1	500	431	86		61	136
		2,4-DCAA	2	500	458	92		61	136
Q2489-02	G4(6-12)	2,4-DCAA	1	500	616	123		61	136
		2,4-DCAA	2	500	659	132		61	136
Q2489-04	G3(6-12)	2,4-DCAA	1	500	404	81		61	136
		2,4-DCAA	2	500	434	87		61	136
Q2489-05	G2(0-6)	2,4-DCAA	1	500	403	81		61	136
		2,4-DCAA	2	500	433	87		61	136
Q2489-06	G2(6-12)	2,4-DCAA	1	500	417	83		61	136
		2,4-DCAA	2	500	437	87		61	136
Q2489-07	G1(0-6)	2,4-DCAA	1	500	407	81		61	136
		2,4-DCAA	2	500	425	85		61	136
Q2489-08	G1(6-12)	2,4-DCAA	1	500	391	78		61	136
		2,4-DCAA	2	500	415	83		61	136
I.BLK-PS031061.D	PIBLK-PS031061.D	2,4-DCAA	1	500	450	90		61	136
		2,4-DCAA	2	500	496	99		61	136
I.BLK-PS031074.D	PIBLK-PS031074.D	2,4-DCAA	1	500	465	93		61	136
		2,4-DCAA	2	500	505	101		61	136
Q2489-03	G3(0-6)	2,4-DCAA	1	500	663	133		61	136
		2,4-DCAA	2	500	488	98		61	136
I.BLK-PS031085.D	PIBLK-PS031085.D	2,4-DCAA	1	500	461	92		61	136
		2,4-DCAA	2	500	487	97		61	136