

**Matrix Spike/Matrix Spike Duplicate Summary**

SW-846

|                 |                                           |                           |                   |
|-----------------|-------------------------------------------|---------------------------|-------------------|
| <b>SDG No.:</b> | <u>Q2489</u>                              | <b>Analytical Method:</b> | <u>8151A</u>      |
| <b>Client:</b>  | <u>Walsh Construction Company II, LLC</u> | <b>DataFile :</b>         | <u>PS031053.D</u> |

|                       | Parameter                         | Sample<br>Spike          | Result            | Result | Units | Rec | Rec<br>Qual | RPD | RPD<br>Qual | Low | Limits<br>High | RPD |
|-----------------------|-----------------------------------|--------------------------|-------------------|--------|-------|-----|-------------|-----|-------------|-----|----------------|-----|
| <b>Lab Sample ID:</b> | <b>Q2489-01MSD<br/>(Column 1)</b> | <b>Client Sample ID:</b> | <b>G4(0-6)MSD</b> |        |       |     |             |     |             |     |                |     |
|                       | 2,4-D                             | 5                        | 0                 | 6.20   | ug/L  | 124 |             | 16  |             | 65  | 135            | 20  |
|                       | 2,4,5-TP(Silvex)                  | 5                        | 0                 | 6.20   | ug/L  | 124 |             | 16  |             | 62  | 139            | 20  |
| <b>Lab Sample ID:</b> | <b>Q2489-01MSD<br/>(Column 2)</b> | <b>Client Sample ID:</b> | <b>G4(0-6)MSD</b> |        |       |     |             |     |             |     |                |     |
|                       | 2,4-D                             | 5                        | 0                 | 5.60   | ug/L  | 112 |             | 27  |             | 65  | 135            | 20  |
|                       | 2,4,5-TP(Silvex)                  | 5                        | 0                 | 6.00   | ug/L  | 120 |             | 4   |             | 62  | 139            | 20  |