



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2489</u>	Analytical Method:	<u>8151A</u>
Client:	<u>Walsh Construction Company II, LLC</u>	Datafile :	<u>PS031049.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB168825BS (Column 1)	2,4-D	5	5.80	ug/L	116				83	130	
	2,4,5-TP(Silvex)	5	5.20	ug/L	104				78	127	
PB168825BS (Column 2)	2,4-D	5	5.00	ug/L	100				83	130	
	2,4,5-TP(Silvex)	5	4.90	ug/L	98				78	127	