

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2489
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/10/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096315.D **Time Analyzed:** 17:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.677	6.578	6.778	55.530	50.000	11.1
4,4'-DDE	6.168	6.069	6.269	51.760	50.000	3.5
4,4'-DDT	6.992	6.892	7.092	51.130	50.000	2.3
Aldrin	5.243	5.146	5.346	52.410	50.000	4.8
alpha-BHC	3.984	3.885	4.085	53.530	50.000	7.1
alpha-Chlordane	5.999	5.900	6.100	52.140	50.000	4.3
beta-BHC	4.498	4.399	4.599	53.160	50.000	6.3
Decachlorobiphenyl	9.021	8.920	9.120	51.270	50.000	2.5
delta-BHC	4.744	4.646	4.846	54.300	50.000	8.6
Dieldrin	6.318	6.219	6.419	52.150	50.000	4.3
Endosulfan I	6.046	5.947	6.147	52.090	50.000	4.2
Endosulfan II	6.757	6.658	6.858	54.660	50.000	9.3
Endosulfan sulfate	7.120	7.019	7.219	53.700	50.000	7.4
Endrin	6.544	6.445	6.645	52.770	50.000	5.5
Endrin aldehyde	6.885	6.786	6.986	59.240	50.000	18.5
Endrin ketone	7.599	7.500	7.700	54.000	50.000	8.0
gamma-BHC (Lindane)	4.312	4.213	4.413	53.440	50.000	6.9
gamma-Chlordane	5.918	5.819	6.019	52.430	50.000	4.9
Heptachlor	4.905	4.806	5.006	54.050	50.000	8.1
Heptachlor epoxide	5.664	5.565	5.765	57.290	50.000	14.6
Methoxychlor	7.464	7.364	7.564	48.930	50.000	-2.1
Tetrachloro-m-xylene	3.536	3.438	3.638	52.140	50.000	4.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2489
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/10/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096315.D **Time Analyzed:** 17:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.858	5.759	5.959	51.400	50.000	2.8
4,4'-DDE	5.305	5.207	5.407	47.200	50.000	-5.6
4,4'-DDT	6.111	6.012	6.212	43.200	50.000	-13.6
Aldrin	4.297	4.200	4.400	47.700	50.000	-4.6
alpha-BHC	3.333	3.235	3.435	47.770	50.000	-4.5
alpha-Chlordane	5.116	5.018	5.218	43.070	50.000	-13.9
beta-BHC	3.961	3.863	4.063	46.890	50.000	-6.2
Decachlorobiphenyl	7.995	7.896	8.096	46.800	50.000	-6.4
delta-BHC	4.194	4.097	4.297	47.540	50.000	-4.9
Dieldrin	5.436	5.338	5.538	47.170	50.000	-5.7
Endosulfan I	5.171	5.073	5.273	40.680	50.000	-18.6
Endosulfan II	6.003	5.904	6.104	46.900	50.000	-6.2
Endosulfan sulfate	6.404	6.306	6.506	46.900	50.000	-6.2
Endrin	5.711	5.612	5.812	44.320	50.000	-11.4
Endrin aldehyde	6.181	6.082	6.282	46.130	50.000	-7.7
Endrin ketone	6.909	6.810	7.010	49.080	50.000	-1.8
gamma-BHC (Lindane)	3.665	3.567	3.767	47.450	50.000	-5.1
gamma-Chlordane	5.052	4.954	5.154	46.970	50.000	-6.1
Heptachlor	4.015	3.917	4.117	46.770	50.000	-6.5
Heptachlor epoxide	4.800	4.702	4.902	46.680	50.000	-6.6
Methoxychlor	6.683	6.584	6.784	43.200	50.000	-13.6
Tetrachloro-m-xylene	2.827	2.729	2.929	46.870	50.000	-6.3