

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2489
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/10/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096325.D **Time Analyzed:** 20:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.677	6.578	6.778	57.540	50.000	15.1
4,4'-DDE	6.169	6.069	6.269	51.940	50.000	3.9
4,4'-DDT	6.991	6.892	7.092	50.370	50.000	0.7
Aldrin	5.246	5.146	5.346	52.630	50.000	5.3
alpha-BHC	3.985	3.885	4.085	53.680	50.000	7.4
alpha-Chlordane	6.000	5.900	6.100	52.700	50.000	5.4
beta-BHC	4.500	4.399	4.599	53.350	50.000	6.7
Decachlorobiphenyl	9.020	8.920	9.120	51.220	50.000	2.4
delta-BHC	4.746	4.646	4.846	54.300	50.000	8.6
Dieldrin	6.319	6.219	6.419	52.450	50.000	4.9
Endosulfan I	6.047	5.947	6.147	52.590	50.000	5.2
Endosulfan II	6.756	6.658	6.858	53.520	50.000	7.0
Endosulfan sulfate	7.120	7.019	7.219	53.590	50.000	7.2
Endrin	6.546	6.445	6.645	52.750	50.000	5.5
Endrin aldehyde	6.885	6.786	6.986	58.600	50.000	17.2
Endrin ketone	7.599	7.500	7.700	54.340	50.000	8.7
gamma-BHC (Lindane)	4.314	4.213	4.413	53.550	50.000	7.1
gamma-Chlordane	5.919	5.819	6.019	53.410	50.000	6.8
Heptachlor	4.906	4.806	5.006	53.850	50.000	7.7
Heptachlor epoxide	5.664	5.565	5.765	55.380	50.000	10.8
Methoxychlor	7.464	7.364	7.564	48.990	50.000	-2.0
Tetrachloro-m-xylene	3.537	3.438	3.638	52.200	50.000	4.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2489
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/10/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096325.D **Time Analyzed:** 20:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.859	5.759	5.959	50.530	50.000	1.1
4,4'-DDE	5.306	5.207	5.407	45.900	50.000	-8.2
4,4'-DDT	6.111	6.012	6.212	41.290	50.000	-17.4
Aldrin	4.299	4.200	4.400	46.480	50.000	-7.0
alpha-BHC	3.335	3.235	3.435	46.640	50.000	-6.7
alpha-Chlordane	5.117	5.018	5.218	42.030	50.000	-15.9
beta-BHC	3.963	3.863	4.063	45.840	50.000	-8.3
Decachlorobiphenyl	7.995	7.896	8.096	46.100	50.000	-7.8
delta-BHC	4.196	4.097	4.297	46.180	50.000	-7.6
Dieldrin	5.436	5.338	5.538	46.520	50.000	-7.0
Endosulfan I	5.171	5.073	5.273	40.080	50.000	-19.8
Endosulfan II	6.003	5.904	6.104	45.810	50.000	-8.4
Endosulfan sulfate	6.405	6.306	6.506	45.960	50.000	-8.1
Endrin	5.711	5.612	5.812	43.130	50.000	-13.7
Endrin aldehyde	6.181	6.082	6.282	45.190	50.000	-9.6
Endrin ketone	6.909	6.810	7.010	49.040	50.000	-1.9
gamma-BHC (Lindane)	3.667	3.567	3.767	46.350	50.000	-7.3
gamma-Chlordane	5.053	4.954	5.154	45.780	50.000	-8.4
Heptachlor	4.016	3.917	4.117	45.590	50.000	-8.8
Heptachlor epoxide	4.801	4.702	4.902	45.530	50.000	-8.9
Methoxychlor	6.684	6.584	6.784	41.670	50.000	-16.7
Tetrachloro-m-xylene	2.829	2.729	2.929	45.930	50.000	-8.1