

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2489
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/11/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096337.D **Time Analyzed:** 00:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.677	6.578	6.778	59.380	50.000	18.8
4,4'-DDE	6.168	6.069	6.269	54.840	50.000	9.7
4,4'-DDT	6.992	6.892	7.092	52.240	50.000	4.5
Aldrin	5.245	5.146	5.346	54.610	50.000	9.2
alpha-BHC	3.985	3.885	4.085	55.800	50.000	11.6
alpha-Chlordane	5.999	5.900	6.100	54.740	50.000	9.5
beta-BHC	4.499	4.399	4.599	54.260	50.000	8.5
Decachlorobiphenyl	9.018	8.920	9.120	59.660	50.000	19.3
delta-BHC	4.745	4.646	4.846	56.370	50.000	12.7
Dieldrin	6.319	6.219	6.419	54.360	50.000	8.7
Endosulfan I	6.046	5.947	6.147	54.890	50.000	9.8
Endosulfan II	6.756	6.658	6.858	55.510	50.000	11.0
Endosulfan sulfate	7.119	7.019	7.219	55.510	50.000	11.0
Endrin	6.545	6.445	6.645	55.040	50.000	10.1
Endrin aldehyde	6.884	6.786	6.986	59.500	50.000	19.0
Endrin ketone	7.599	7.500	7.700	56.600	50.000	13.2
gamma-BHC (Lindane)	4.313	4.213	4.413	55.290	50.000	10.6
gamma-Chlordane	5.918	5.819	6.019	54.930	50.000	9.9
Heptachlor	4.905	4.806	5.006	56.180	50.000	12.4
Heptachlor epoxide	5.663	5.565	5.765	57.590	50.000	15.2
Methoxychlor	7.464	7.364	7.564	50.840	50.000	1.7
Tetrachloro-m-xylene	3.537	3.438	3.638	54.090	50.000	8.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2489
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/11/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096337.D **Time Analyzed:** 00:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.859	5.759	5.959	51.150	50.000	2.3
4,4'-DDE	5.306	5.207	5.407	47.110	50.000	-5.8
4,4'-DDT	6.111	6.012	6.212	42.350	50.000	-15.3
Aldrin	4.299	4.200	4.400	47.260	50.000	-5.5
alpha-BHC	3.334	3.235	3.435	47.220	50.000	-5.6
alpha-Chlordane	5.117	5.018	5.218	42.870	50.000	-14.3
beta-BHC	3.962	3.863	4.063	46.160	50.000	-7.7
Decachlorobiphenyl	7.995	7.896	8.096	52.900	50.000	5.8
delta-BHC	4.196	4.097	4.297	47.080	50.000	-5.8
Dieldrin	5.435	5.338	5.538	47.230	50.000	-5.5
Endosulfan I	5.172	5.073	5.273	40.440	50.000	-19.1
Endosulfan II	6.003	5.904	6.104	46.790	50.000	-6.4
Endosulfan sulfate	6.405	6.306	6.506	47.100	50.000	-5.8
Endrin	5.711	5.612	5.812	44.300	50.000	-11.4
Endrin aldehyde	6.181	6.082	6.282	45.740	50.000	-8.5
Endrin ketone	6.909	6.810	7.010	51.240	50.000	2.5
gamma-BHC (Lindane)	3.666	3.567	3.767	46.850	50.000	-6.3
gamma-Chlordane	5.053	4.954	5.154	46.610	50.000	-6.8
Heptachlor	4.016	3.917	4.117	46.090	50.000	-7.8
Heptachlor epoxide	4.801	4.702	4.902	46.570	50.000	-6.9
Methoxychlor	6.684	6.584	6.784	42.670	50.000	-14.7
Tetrachloro-m-xylene	2.829	2.729	2.929	46.390	50.000	-7.2