

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2489
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/11/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096341.D **Time Analyzed:** 15:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.677	6.578	6.778	56.770	50.000	13.5
4,4'-DDE	6.169	6.069	6.269	54.720	50.000	9.4
4,4'-DDT	6.991	6.892	7.092	57.480	50.000	15.0
Aldrin	5.245	5.146	5.346	54.540	50.000	9.1
alpha-BHC	3.984	3.885	4.085	55.370	50.000	10.7
alpha-Chlordane	5.998	5.900	6.100	55.040	50.000	10.1
beta-BHC	4.498	4.399	4.599	54.270	50.000	8.5
Decachlorobiphenyl	9.018	8.920	9.120	54.220	50.000	8.4
delta-BHC	4.744	4.646	4.846	53.960	50.000	7.9
Dieldrin	6.318	6.219	6.419	53.720	50.000	7.4
Endosulfan I	6.045	5.947	6.147	54.460	50.000	8.9
Endosulfan II	6.755	6.658	6.858	57.930	50.000	15.9
Endosulfan sulfate	7.119	7.019	7.219	55.140	50.000	10.3
Endrin	6.544	6.445	6.645	58.540	50.000	17.1
Endrin aldehyde	6.884	6.786	6.986	59.740	50.000	19.5
Endrin ketone	7.599	7.500	7.700	54.480	50.000	9.0
gamma-BHC (Lindane)	4.312	4.213	4.413	55.350	50.000	10.7
gamma-Chlordane	5.918	5.819	6.019	53.740	50.000	7.5
Heptachlor	4.905	4.806	5.006	56.270	50.000	12.5
Heptachlor epoxide	5.664	5.565	5.765	57.970	50.000	15.9
Methoxychlor	7.463	7.364	7.564	56.040	50.000	12.1
Tetrachloro-m-xylene	3.536	3.438	3.638	53.660	50.000	7.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2489
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/11/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096341.D **Time Analyzed:** 15:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.858	5.759	5.959	49.880	50.000	-0.2
4,4'-DDE	5.305	5.207	5.407	47.260	50.000	-5.5
4,4'-DDT	6.110	6.012	6.212	47.590	50.000	-4.8
Aldrin	4.298	4.200	4.400	47.250	50.000	-5.5
alpha-BHC	3.333	3.235	3.435	46.930	50.000	-6.1
alpha-Chlordane	5.116	5.018	5.218	43.220	50.000	-13.6
beta-BHC	3.961	3.863	4.063	45.790	50.000	-8.4
Decachlorobiphenyl	7.994	7.896	8.096	48.780	50.000	-2.4
delta-BHC	4.195	4.097	4.297	46.630	50.000	-6.7
Dieldrin	5.434	5.338	5.538	47.260	50.000	-5.5
Endosulfan I	5.171	5.073	5.273	40.870	50.000	-18.3
Endosulfan II	6.002	5.904	6.104	47.290	50.000	-5.4
Endosulfan sulfate	6.404	6.306	6.506	47.140	50.000	-5.7
Endrin	5.710	5.612	5.812	47.350	50.000	-5.3
Endrin aldehyde	6.180	6.082	6.282	46.840	50.000	-6.3
Endrin ketone	6.908	6.810	7.010	50.430	50.000	0.9
gamma-BHC (Lindane)	3.665	3.567	3.767	46.770	50.000	-6.5
gamma-Chlordane	5.052	4.954	5.154	46.750	50.000	-6.5
Heptachlor	4.015	3.917	4.117	47.690	50.000	-4.6
Heptachlor epoxide	4.800	4.702	4.902	46.550	50.000	-6.9
Methoxychlor	6.683	6.584	6.784	47.910	50.000	-4.2
Tetrachloro-m-xylene	2.827	2.729	2.929	45.670	50.000	-8.7