

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2489
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/11/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096346.D **Time Analyzed:** 16:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.677	6.578	6.778	57.200	50.000	14.4
4,4'-DDE	6.169	6.069	6.269	53.900	50.000	7.8
4,4'-DDT	6.992	6.892	7.092	56.960	50.000	13.9
Aldrin	5.245	5.146	5.346	54.030	50.000	8.1
alpha-BHC	3.984	3.885	4.085	55.110	50.000	10.2
alpha-Chlordane	5.999	5.900	6.100	54.140	50.000	8.3
beta-BHC	4.498	4.399	4.599	53.620	50.000	7.2
Decachlorobiphenyl	9.019	8.920	9.120	53.390	50.000	6.8
delta-BHC	4.745	4.646	4.846	53.810	50.000	7.6
Dieldrin	6.318	6.219	6.419	53.830	50.000	7.7
Endosulfan I	6.046	5.947	6.147	53.940	50.000	7.9
Endosulfan II	6.756	6.658	6.858	56.580	50.000	13.2
Endosulfan sulfate	7.119	7.019	7.219	54.760	50.000	9.5
Endrin	6.545	6.445	6.645	57.950	50.000	15.9
Endrin aldehyde	6.885	6.786	6.986	59.310	50.000	18.6
Endrin ketone	7.599	7.500	7.700	53.820	50.000	7.6
gamma-BHC (Lindane)	4.312	4.213	4.413	54.970	50.000	9.9
gamma-Chlordane	5.918	5.819	6.019	53.260	50.000	6.5
Heptachlor	4.905	4.806	5.006	56.230	50.000	12.5
Heptachlor epoxide	5.664	5.565	5.765	57.750	50.000	15.5
Methoxychlor	7.464	7.364	7.564	55.820	50.000	11.6
Tetrachloro-m-xylene	3.536	3.438	3.638	53.430	50.000	6.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2489
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/11/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096346.D **Time Analyzed:** 16:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.858	5.759	5.959	49.470	50.000	-1.1
4,4'-DDE	5.305	5.207	5.407	46.570	50.000	-6.9
4,4'-DDT	6.110	6.012	6.212	46.300	50.000	-7.4
Aldrin	4.298	4.200	4.400	46.760	50.000	-6.5
alpha-BHC	3.333	3.235	3.435	46.620	50.000	-6.8
alpha-Chlordane	5.116	5.018	5.218	42.850	50.000	-14.3
beta-BHC	3.962	3.863	4.063	45.560	50.000	-8.9
Decachlorobiphenyl	7.994	7.896	8.096	47.950	50.000	-4.1
delta-BHC	4.195	4.097	4.297	46.370	50.000	-7.3
Dieldrin	5.435	5.338	5.538	47.280	50.000	-5.4
Endosulfan I	5.171	5.073	5.273	40.200	50.000	-19.6
Endosulfan II	6.002	5.904	6.104	46.510	50.000	-7.0
Endosulfan sulfate	6.404	6.306	6.506	46.610	50.000	-6.8
Endrin	5.710	5.612	5.812	46.680	50.000	-6.6
Endrin aldehyde	6.180	6.082	6.282	46.400	50.000	-7.2
Endrin ketone	6.908	6.810	7.010	50.500	50.000	1.0
gamma-BHC (Lindane)	3.666	3.567	3.767	46.430	50.000	-7.1
gamma-Chlordane	5.052	4.954	5.154	46.540	50.000	-6.9
Heptachlor	4.015	3.917	4.117	47.130	50.000	-5.7
Heptachlor epoxide	4.800	4.702	4.902	45.990	50.000	-8.0
Methoxychlor	6.683	6.584	6.784	47.230	50.000	-5.5
Tetrachloro-m-xylene	2.827	2.729	2.929	45.460	50.000	-9.1