

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2489

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 07/07/2025 07/07/2025

Client Sample No. (PEM): PEM - PL096238.D Date Analyzed: 07/07/2025

Lab Sample No.(PEM): PEM Time Analyzed: 10:26

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.020	8.920	9.120	21.820	20.000	9.1
Tetrachloro-m-xylene	3.537	3.490	3.590	20.640	20.000	3.2
alpha-BHC	3.985	3.930	4.040	10.150	10.000	1.5
beta-BHC	4.500	4.450	4.550	10.650	10.000	6.5
gamma-BHC (Lindane)	4.313	4.260	4.360	10.640	10.000	6.4
Endrin	6.546	6.480	6.620	55.200	50.000	10.4
4,4'-DDT	6.992	6.920	7.060	106.050	100.000	6.1
Methoxychlor	7.465	7.390	7.540	251.880	250.000	0.8

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 07/07/2025 07/07/2025

Client Sample No. (PEM): PEM - PL096238.D Date Analyzed: 07/07/2025

Lab Sample No.(PEM): PEM Time Analyzed: 10:26

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.996	7.900	8.100	21.970	20.000	9.9
Tetrachloro-m-xylene	2.829	2.780	2.880	21.010	20.000	5.1
alpha-BHC	3.335	3.280	3.390	10.410	10.000	4.1
beta-BHC	3.963	3.910	4.010	10.960	10.000	9.6
gamma-BHC (Lindane)	3.667	3.620	3.720	10.560	10.000	5.6
Endrin	5.713	5.640	5.780	55.380	50.000	10.8
4,4'-DDT	6.113	6.040	6.180	113.720	100.000	13.7
Methoxychlor	6.685	6.610	6.760	248.840	250.000	-0.5